



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB100010

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.00	12/21/2018	17:20
2	CAL2	1	Water			22.2	7.00	12/21/2018	17:22
3	CAL3	1	Water			22.3	10.03	12/21/2018	17:24
4	ICV	1	Water			22.2	7.01	12/21/2018	17:26
5	CCV1	1	Water			22.3	2.00	12/21/2018	17:28
6	J6533-08	1	Water			18.7	6.73	12/21/2018	17:30
7	J6533-08DUP	1	Water			18.8	6.74	12/21/2018	17:32
8	J6533-11	1	Water			18.2	6.82	12/21/2018	17:34
9	J6533-14	1	Water			18.6	5.89	12/21/2018	17:36
10	J6533-17	1	Water			19.3	5.22	12/21/2018	17:38
11	J6533-20	1	Water			20.3	5.47	12/21/2018	17:40
12	CCV2	1	Water			22.2	12.00	12/21/2018	17:42

WORKLIST(Hardcopy Internal Chain)

LB 100010

WorkList Name : corrosivity122118-1

WorkList ID : 120728

Date : 12/21/2018 3:21:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/21/2018	Water	J6533-08	Corrosivity	Cool 4 deg C	PSEG01	E52	COMP-1	12/21/2018	9040C
12/21/2018	Water	J6533-11	Corrosivity	Cool 4 deg C	PSEG01	E52	COMP-2	12/21/2018	9040C
12/21/2018	Water	J6533-14	Corrosivity	Cool 4 deg C	PSEG01	E52	COMP-3	12/21/2018	9040C
12/21/2018	Water	J6533-17	Corrosivity	Cool 4 deg C	PSEG01	E52	COMP-4	12/21/2018	9040C
12/21/2018	Water	J6533-20	Corrosivity	Cool 4 deg C	PSEG01	E52	COMP-5	12/21/2018	9040C

Date/Time 12/21/18 5:05
 Received by: SP
 Relinquished by: CP

Date/Time 12/21/18 5:45
 Received by: SP
 Relinquished by: SP