



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB100012

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	12/21/2018	17:20
2	CAL2	1	Water			22.2	7.00	12/21/2018	17:22
3	CAL3	1	Water			22.3	10.03	12/21/2018	17:24
4	ICV	1	Water			22.2	7.01	12/21/2018	17:26
5	CCV1	1	Water			22.1	2.00	12/21/2018	17:46
6	J6188-06	1	Solid	20.02	20	22.1	8.47	12/21/2018	17:48
7	J6188-06DUP	1	Solid	20.03	20	22.2	8.48	12/21/2018	17:50
8	J6498-07	1	Solid	20.02	20	18.5	9.57	12/21/2018	17:52
9	J6498-08	1	Solid	20.02	20	17.3	10.20	12/21/2018	17:54
10	J6505-02	1	Solid	20.04	20	22.9	6.57	12/21/2018	17:56
11	J6505-04	1	Solid	20.04	20	22.8	7.55	12/21/2018	17:58
12	J6505-06	1	Solid	20.02	20	22.9	8.12	12/21/2018	18:00
13	J6505-08	1	Solid	20.02	20	22.9	7.11	12/21/2018	18:02
14	J6505-10	1	Solid	20.03	20	22.8	7.65	12/21/2018	18:04
15	J6505-12	1	Solid	20.02	20	22.7	8.12	12/21/2018	18:06
16	CCV2	1	Water			22.3	12.00	12/21/2018	18:08
17	J6505-14	1	Solid	20.02	20	22.7	8.15	12/21/2018	18:10
18	J6505-16	1	Solid	20.03	20	22.8	8.19	12/21/2018	18:12
19	J6505-16DUP	1	Solid	20.04	20	22.9	8.20	12/21/2018	18:14
20	J6505-18	1	Solid	20.02	20	22.6	7.14	12/21/2018	18:16

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			22.2	2.00	12/21/2018	18:18

WORKLIST(Hardcopy Internal Chain)

12/21/2018

WorkList Name : corrosivity122118

WorkList ID : 120727

Date : 12/21/2018 3:21:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/21/2018	Solid	J6188-06	Corrosivity	Cool 4 deg C	PSEG05		BU-03-122118	12/21/2018	9045D
12/19/2018	Solid	J6498-07	Corrosivity	Cool 4 deg C	USAR03	D31	SSSI-1219-S-TCLP-5	12/19/2018	9045D
12/19/2018	Solid	J6498-08	Corrosivity	Cool 4 deg C	USAR03	D31	SSSI-1219-S-TCLP-6	12/19/2018	9045D
12/18/2018	Solid	J6505-02	Corrosivity	Cool 4 deg C	TULL02	D22	SB-57(0-12)-COMP	12/18/2018	9045D
12/18/2018	Solid	J6505-04	Corrosivity	Cool 4 deg C	TULL02	D22	SB-58(0-23)-COMP	12/18/2018	9045D
12/18/2018	Solid	J6505-06	Corrosivity	Cool 4 deg C	TULL02	D22	SB-59(0-23)-COMP	12/18/2018	9045D
12/19/2018	Solid	J6505-08	Corrosivity	Cool 4 deg C	TULL02	D22	SB-60(0-23)-COMP	12/19/2018	9045D
12/19/2018	Solid	J6505-10	Corrosivity	Cool 4 deg C	TULL02	D22	SB-61(0-23)-COMP	12/19/2018	9045D
12/19/2018	Solid	J6505-12	Corrosivity	Cool 4 deg C	TULL02	D22	SB-62(0-23)-COMP	12/19/2018	9045D
12/19/2018	Solid	J6505-14	Corrosivity	Cool 4 deg C	TULL02	D22	SB-63(0-23)-COMP	12/19/2018	9045D
12/19/2018	Solid	J6505-16	Corrosivity	Cool 4 deg C	TULL02	D22	SB-64(0-23)-COMP	12/19/2018	9045D
12/19/2018	Solid	J6505-18	Corrosivity	Cool 4 deg C	TULL02	D22	SB-65(0-12)-COMP	12/19/2018	9045D

Date/Time 12/21/18 5:05

Received by: SP

Relinquished by: SP

Date/Time 12/21/18 6:30

Received by: CP

Relinquished by: SP