

LB 100026


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53228



A= 0.0000e+000

B= 1.8601e-004

C= 4.1128e-003

Rho= 0.9993376

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-33	0.000	-33				
0.2	0.200	0.194	-0.006	1023	0.0 %	1023				
2.5	2.500	2.589	0.089	13895	0.0 %	13895				
5.0	5.000	4.793	-0.207	25747	0.0 %	25747				
7.5	7.500	7.720	0.220	41483	0.0 %	41483				
10.0	10.000	9.905	-0.095	53228	0.0 %	53228				

LB100026 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Dec 2018 10:38:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	-33	-	0.0000	7471B	PPB	26 Dec 2018 10:43:56	S	Std
0.2 - 1	SC2	1023	-	0.2000	7471B	PPB	26 Dec 2018 10:46:13	S	Std
2.5 - 1	SC2.5	13895	-	2.5000	7471B	PPB	26 Dec 2018 10:48:32	S	Std
5.0 - 1	SC5	25747	-	5.0000	7471B	PPB	26 Dec 2018 10:50:49	S	Std
7.5 - 1	SC7.5	41483	-	7.5000	7471B	PPB	26 Dec 2018 10:53:06	S	Std
10.0 - 1	SC10	53228	-	10.0000	7471B	PPB	26 Dec 2018 10:58:40	S	Std
ICV40 - 1	ICV40	22447	4.1795	-7471B	PPB	PPB	26 Dec 2018 11:01:34	U	SMPL
ICB40 - 1	ICB40	-72	-0.0093	-7471B	PPB	PPB	26 Dec 2018 11:03:49	U	SMPL
CCV83 - 1	CCV83	25848	4.8121	-7471B	PPB	PPB	26 Dec 2018 11:06:06	U	SMPL
CCB83 - 1	CCB83	-75	-0.0098	-7471B	PPB	PPB	26 Dec 2018 11:08:22	U	SMPL
CRA - 1	CRA	1054	0.2002	-7471B	PPB	PPB	26 Dec 2018 11:10:39	U	SMPL
HighStd - 1	HighStd	54254	10.0960	-7471B	PPB	PPB	26 Dec 2018 11:12:55	U	SMPL
ChkStd - 1	ChkStd	38786	7.2187	-7471B	PPB	PPB	26 Dec 2018 11:15:11	U	SMPL
PB115973BL - 1	PBS	2	0.0045	-7471B	PPB	PPB	26 Dec 2018 11:17:29	U	SMPL
PB115973BS - 1	LCSS	18691	3.4808	-7471B	PPB	PPB	26 Dec 2018 11:19:48	U	SMPL
J6188-05 - 1	BU-03-122118	1193	0.2260	-7471B	PPB	PPB	26 Dec 2018 11:22:06	U	SMPL
J6498-02 - 1	SSSI-1218-S-DH-9-T	2908	0.5450	-7471B	PPB	PPB	26 Dec 2018 11:24:23	U	SMPL
J6498-04 - 1	SSSI-1218-S-DH-9-B	2113	0.3972	-7471B	PPB	PPB	26 Dec 2018 11:26:40	U	SMPL
J6498-04DUP - 1	SSSI-1218-S-DH-9-BDUP	2031	0.3819	-7471B	PPB	PPB	26 Dec 2018 11:28:58	U	SMPL
J6498-04MS - 1	SSSI-1218-S-DH-9-BMS	21723	4.0448	-7471B	PPB	PPB	26 Dec 2018 11:31:15	U	SMPL
CCV84 - 1	CCV84	25602	4.7664	-7471B	PPB	PPB	26 Dec 2018 11:33:31	U	SMPL
CCB84 - 1	CCB84	-70	-0.0089	-7471B	PPB	PPB	26 Dec 2018 11:35:47	U	SMPL
J6498-04MSD - 1	SSSI-1218-S-DH-9-BMSD	22031	4.1021	-7471B	PPB	PPB	26 Dec 2018 11:38:04	U	SMPL
J6498-06 - 1	SSSI-1218-S-AB-6	5320	0.9937	-7471B	PPB	PPB	26 Dec 2018 11:40:20	U	SMPL
J6505-01 - 1	SB-57(0-12)-COMP	3512	0.6574	-7471B	PPB	PPB	26 Dec 2018 11:42:37	U	SMPL
J6505-03 - 1	SB-58(0-23)-COMP	2418	0.4539	-7471B	PPB	PPB	26 Dec 2018 11:44:54	U	SMPL
J6505-05 - 1	SB-59(0-23)-COMP	72625	13.5132	-7471B	PPB	PPB	26 Dec 2018 11:47:11	U	SMPL
J6505-07 - 1	SB-60(0-23)-COMP	1767	0.3328	-7471B	PPB	PPB	26 Dec 2018 11:49:28	U	SMPL
J6505-09 - 1	SB-61(0-23)-COMP	11167	2.0813	-7471B	PPB	PPB	26 Dec 2018 11:51:48	U	SMPL
J6505-11 - 1	SB-62(0-23)-COMP	5540	1.0346	-7471B	PPB	PPB	26 Dec 2018 11:54:06	U	SMPL
J6505-13 - 1	SB-63(0-23)-COMP	2059	0.3871	-7471B	PPB	PPB	26 Dec 2018 11:56:24	U	SMPL
J6505-15 - 1	SB-64(0-23)-COMP	478	0.0930	-7471B	PPB	PPB	26 Dec 2018 11:58:40	U	SMPL
CCV85 - 1	CCV85	25870	4.8162	-7471B	PPB	PPB	26 Dec 2018 12:00:56	U	SMPL
CCB85 - 1	CCB85	-86	-0.0119	-7471B	PPB	PPB	26 Dec 2018 12:03:13	U	SMPL
J6505-17 - 1	SB-65(0-12)-COMP	680	0.1306	-7471B	PPB	PPB	26 Dec 2018 12:05:30	U	SMPL
J6498-04LX5 - 1		393	0.0772	-7471B	PPB	PPB	26 Dec 2018 12:07:46	U	SMPL
J6498-04A - 1		24360	4.5353	-7471B	PPB	PPB	26 Dec 2018 12:10:03	U	SMPL
J6505-05DLX2 - 1		36621	6.8160	-7471B	PPB	PPB	26 Dec 2018 12:12:20	U	SMPL
CCV86 - 1	CCV86	26072	4.8538	-7471B	PPB	PPB	26 Dec 2018 12:14:37	U	SMPL
CCB86 - 1	CCB86	-56	-0.0063	-7471B	PPB	PPB	26 Dec 2018 12:16:56	U	SMPL