

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB100039

ANALYST: sweta
 SUPERVISOR REVIEW BY: jignesh
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP69572
IODINE SOLUTION .025N 1L	W2410
Starch Solution, 4L	W2277

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB115976BL		1	5.05	50	2.00	0.00	1.90	1.90	0.10	0.00	0.00	12/26/2018	11:15
2	J6188-06		1	5.03	50	2.00	1.90	3.70	1.80	0.20	0.10	7.95	12/26/2018	11:17
3	J6188-06DUP		1	5.04	50	2.00	3.70	5.52	1.82	0.18	0.08	6.35	12/26/2018	11:19

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume