



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB100041

Analyst By : sweta

Supervisor Review By : jignesh

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	12/26/2018	15:45
2	CAL2	1	Water			21.2	7.00	12/26/2018	15:47
3	CAL3	1	Water			21.3	10.04	12/26/2018	15:49
4	ICV	1	Water			21.1	7.01	12/26/2018	15:51
5	CCV1	1	Water			21.4	2.00	12/26/2018	15:53
6	J6575-02	1	Water			13.6	6.60	12/26/2018	16:07
7	J6575-02DUP	1	Water			13.7	6.61	12/26/2018	16:09
8	CCV2	1	Water			21.3	12.00	12/26/2018	16:11

WORKLIST(Hardcopy Internal Chain)

UB 400041

WorkList Name : corrosivity122618-1

WorkList ID : 120784

Date : 12/26/2018 3:33:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2018	Water	J6575-02	Corrosivity	Cool 4 deg C	PSEG01	E31	SS-122918-A	12/26/2018	9040C

Date/Time 12/26/18 3:35
 Received by: SP
 Relinquished by: CP

Date/Time 12/26/18 4:15
 Received by: CP
 Relinquished by: SP