



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB100042

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	12/26/2018	15:45
2	CAL2	1	Water			21.2	7.00	12/26/2018	15:47
3	CAL3	1	Water			21.3	10.04	12/26/2018	15:49
4	ICV	1	Water			21.1	7.01	12/26/2018	15:51
5	CCV1	1	Water			21.4	2.00	12/26/2018	15:53
6	J6571-03	1	Water			13.9	5.97	12/26/2018	15:55
7	J6571-04	1	Water			15.2	5.37	12/26/2018	15:57
8	J6571-06	1	Water			16.6	8.83	12/26/2018	15:59
9	J6571-07	1	Water			15.5	8.86	12/26/2018	16:01
10	J6571-08	1	Water			16.7	8.72	12/26/2018	16:03
11	J6571-09	1	Water			17.0	9.05	12/26/2018	16:05
12	J6575-01	1	Water			13.6	6.60	12/26/2018	16:07
13	J6575-01DUP	1	Water			13.7	6.61	12/26/2018	16:09
14	CCV2	1	Water			21.3	12.00	12/26/2018	16:11

WORKLIST(Hardcopy Internal Chain)

VB 106042

WorkList Name : pH122618

WorkList ID : 120785

Date : 12/26/2018 3:33:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2018	Water	J6571-03	pH	Cool 4 deg C	PSEG01	E31	30061	12/26/2018	9040C
12/26/2018	Water	J6571-04	pH	Cool 4 deg C	PSEG01	E31	50082-50083	12/26/2018	9040C
12/26/2018	Water	J6571-06	pH	Cool 4 deg C	PSEG01	E31	50084	12/26/2018	9040C
12/26/2018	Water	J6571-07	pH	Cool 4 deg C	PSEG01	E31	50085	12/26/2018	9040C
12/26/2018	Water	J6571-08	pH	Cool 4 deg C	PSEG01	E31	50086	12/26/2018	9040C
12/26/2018	Water	J6571-09	pH	Cool 4 deg C	PSEG01	E31	50087	12/26/2018	9040C
12/26/2018	Water	J6575-01	pH	Cool 4 deg C	PSEG01	E31	SS-122918-A	12/26/2018	9040C

Date/Time 12/26/18 3:35

Received by: SP

Relinquished by: CP

Date/Time 12/26/18 4:15

Received by: CP

Relinquished by: SP