



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB100043

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	12/26/2018	15:45
2	CAL2	1	Water			21.2	7.00	12/26/2018	15:47
3	CAL3	1	Water			21.3	10.04	12/26/2018	15:49
4	ICV	1	Water			21.1	7.01	12/26/2018	15:51
5	CCV1	1	Water			21.2	2.00	12/26/2018	16:13
6	J6571-01	1	Solid	20.02	20	20.3	6.20	12/26/2018	16:15
7	J6571-01DUP	1	Solid	20.03	20	20.4	6.21	12/26/2018	16:17
8	J6571-02	1	Solid	20.02	20	20.5	7.71	12/26/2018	16:19
9	CCV2	1	Water			21.3	12.00	12/26/2018	16:21

WORKLIST(Hardcopy Internal Chain)

VB100043

WorkList Name : pH122618-1

WorkList ID : 120786

Date : 12/26/2018 3:35:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2018	Solid	J6571-01	pH	Cool 4 deg C	PSEG01	E31	50072	12/26/2018	9045D
12/26/2018	Solid	J6571-02	pH	Cool 4 deg C	PSEG01	E31	10827	12/26/2018	9045D

Date/Time 12/26/18 3:37

Received by: SP

Relinquished by: CP

Date/Time 12/26/18 4:25

Received by: CP

Relinquished by: SP