

7470A


LB 100052

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38927



A= 0.0000e+000

B= 2.5712e-004

C= 2.1151e-002

Rho= 0.9998844

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	79	0.000	79				
0.2	0.200	0.226	0.026	797	0.0 %	797				
2.5	2.500	2.476	-0.024	9548	0.0 %	9548				
5.0	5.000	4.885	-0.115	18916	0.0 %	18916				
7.5	7.500	7.542	0.042	29250	0.0 %	29250				
10.0	10.000	10.030	0.030	38927	0.0 %	38927	0			

LB100052 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Dec 2018 09:42:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.00007470A	PPB		27 Dec 2018 10:15:07	S	Std
0.2 - 1	50.2	797	-	0.20007470A	PPB		27 Dec 2018 10:17:16	S	Std
2.5 - 1	52.5	9548	-	2.50007470A	PPB		27 Dec 2018 10:19:25	S	Std
5.0 - 1	55	18916	-	5.00007470A	PPB		27 Dec 2018 10:21:33	S	Std
7.5 - 1	57.5	29250	-	7.50007470A	PPB		27 Dec 2018 10:23:41	S	Std
10.0 - 1	510	38927	-	10.00007470A	PPB		27 Dec 2018 10:25:48	S	Std
ICV42 - 1	ICV42	14904	3.8532	-7470A	PPB		27 Dec 2018 10:28:30	U	SMPL
ICB42 - 1	ICB42	-93	-0.0028	-7470A	PPB		27 Dec 2018 10:30:36	U	SMPL
CCV92 - 1	CCV92	18771	4.8475	-7470A	PPB		27 Dec 2018 10:32:42	U	SMPL
CCB92 - 1	CCB92	-98	-0.0040	-7470A	PPB		27 Dec 2018 10:34:49	U	SMPL
CRA - 1	CRA	781	0.2220	-7470A	PPB		27 Dec 2018 10:36:56	U	SMPL
HighStd - 1	HighStd	38178	9.8373	-7470A	PPB		27 Dec 2018 10:39:02	U	SMPL
ChkStd - 1	ChkStd	27106	6.9905	-7470A	PPB		27 Dec 2018 10:41:09	U	SMPL
PB115964BL - 1	PBW	167	0.0641	-7470A	PPB		27 Dec 2018 10:43:17	U	SMPL
PB115964BS - 1	LCSW	13328	3.4480	-7470A	PPB		27 Dec 2018 10:49:58	U	SMPL
J6188-06 - 1	BU-03-122118	234	0.0813	-7470A	PPB		27 Dec 2018 10:52:05	U	SMPL
J6188-06DUP - 1	BU-03-122118DUP	252	0.0859	-7470A	PPB		27 Dec 2018 10:54:13	U	SMPL
J6188-06MS - 1	BU-03-122118MS	13953	3.6087	-7470A	PPB		27 Dec 2018 10:56:21	U	SMPL
J6188-06MSD - 1	BU-03-122118MSD	14427	3.7306	-7470A	PPB		27 Dec 2018 10:58:30	U	SMPL
J6505-02 - 1	SB-57(0-12)-COMP	682	0.1965	-7470A	PPB		27 Dec 2018 11:00:37	U	SMPL
CCV93 - 1	CCV93	18662	4.8194	-7470A	PPB		27 Dec 2018 11:02:44	U	SMPL
CCB93 - 1	CCB93	-81	0.0003	-7470A	PPB		27 Dec 2018 11:04:51	U	SMPL
J6505-04 - 1	SB-58(0-23)-COMP	715	0.2050	-7470A	PPB		27 Dec 2018 11:06:57	U	SMPL
J6505-06 - 1	SB-59(0-23)-COMP	513	0.1531	-7470A	PPB		27 Dec 2018 11:09:05	U	SMPL
J6505-08 - 1	SB-60(0-23)-COMP	167	0.0641	-7470A	PPB		27 Dec 2018 11:11:12	U	SMPL
J6505-10 - 1	SB-61(0-23)-COMP	2	0.0217	-7470A	PPB		27 Dec 2018 11:13:19	U	SMPL
J6505-12 - 1	SB-62(0-23)-COMP	268	0.0901	-7470A	PPB		27 Dec 2018 11:15:27	U	SMPL
J6505-14 - 1	SB-63(0-23)-COMP	267	0.0898	-7470A	PPB		27 Dec 2018 11:17:34	U	SMPL
J6505-16 - 1	SB-64(0-23)-COMP	266	0.0895	-7470A	PPB		27 Dec 2018 11:19:43	U	SMPL
J6505-18 - 1	SB-65(0-12)-COMP	334	0.1070	-7470A	PPB		27 Dec 2018 11:21:51	U	SMPL
J6533-01 - 1	22949	466	0.1410	-7470A	PPB		27 Dec 2018 11:24:00	U	SMPL
J6544-01 - 1	23393	328	0.1055	-7470A	PPB		27 Dec 2018 11:26:07	U	SMPL
CCV94 - 1	CCV94	18706	4.8307	-7470A	PPB		27 Dec 2018 11:28:14	U	SMPL
CCB94 - 1	CCB94	-102	-0.0051	-7470A	PPB		27 Dec 2018 11:30:21	U	SMPL
J6548-19 - 1	G-A	155	0.0610	-7470A	PPB		27 Dec 2018 11:32:28	U	SMPL
J6548-20 - 1	G-B	473	0.1428	-7470A	PPB		27 Dec 2018 11:34:35	U	SMPL
J6548-21 - 1	G-C	381	0.1191	-7470A	PPB		27 Dec 2018 11:36:42	U	SMPL
PB115965BS - 1	LCSW	14281	3.6930	-7470A	PPB		27 Dec 2018 11:40:57	U	SMPL
PB115965BL - 1	PBW	-75	0.0019	-7470A	PPB		27 Dec 2018 11:44:04	U	SMPL
J6533-08 - 1	COMP-1	182	0.0679	-7470A	PPB		27 Dec 2018 11:46:30	U	SMPL
J6533-08DUP - 1	COMP-1DUP	173	0.0656	-7470A	PPB		27 Dec 2018 11:48:38	U	SMPL
J6533-08MS - 1	COMP-1MS	14471	3.7419	-7470A	PPB		27 Dec 2018 11:50:46	U	SMPL
J6533-08MSD - 1	COMP-1MSD	14732	3.8090	-7470A	PPB		27 Dec 2018 11:52:55	U	SMPL
J6533-11 - 1	COMP-2	176	0.0664	-7470A	PPB		27 Dec 2018 11:55:02	U	SMPL
CCV95 - 1	CCV95	19041	4.9169	-7470A	PPB		27 Dec 2018 11:57:09	U	SMPL
CCB95 - 1	CCB95	9	0.0235	-7470A	PPB		27 Dec 2018 11:59:17	U	SMPL
J6533-14 - 1	COMP-3	232	0.0808	-7470A	PPB		27 Dec 2018 12:01:24	U	SMPL
J6533-17 - 1	COMP-4	258	0.0875	-7470A	PPB		27 Dec 2018 12:03:31	U	SMPL
J6533-20 - 1	COMP-5	216	0.0767	-7470A	PPB		27 Dec 2018 12:05:38	U	SMPL
PB115966BL - 1	PBW	-36	0.0119	-7470A	PPB		27 Dec 2018 12:07:46	U	SMPL
PB115966BS - 1	LCSW	14448	3.7360	-7470A	PPB		27 Dec 2018 12:09:53	U	SMPL
J6548-01 - 1	D-1	99	0.0466	-7470A	PPB		27 Dec 2018 12:12:01	U	SMPL
J6548-01DUP - 1	D-1DUP	119	0.0517	-7470A	PPB		27 Dec 2018 12:14:09	U	SMPL
J6548-01MS - 1	D-1MS	14358	3.7128	-7470A	PPB		27 Dec 2018 12:16:17	U	SMPL
J6548-01MSD - 1	D-1MSD	14590	3.7725	-7470A	PPB		27 Dec 2018 12:22:18	U	SMPL
J6548-02 - 1	D-2	396	0.1230	-7470A	PPB		27 Dec 2018 12:24:24	U	SMPL
CCV96 - 1	CCV96	18806	4.8565	-7470A	PPB		27 Dec 2018 12:26:32	U	SMPL
CCB96 - 1	CCB96	2	0.0217	-7470A	PPB		27 Dec 2018 12:28:39	U	SMPL
J6548-03 - 1	D-3	392	0.1219	-7470A	PPB		27 Dec 2018 12:30:46	U	SMPL
J6548-04 - 1	D-4	401	0.1243	-7470A	PPB		27 Dec 2018 12:32:53	U	SMPL
J6548-05 - 1	D-5	402	0.1245	-7470A	PPB		27 Dec 2018 12:35:00	U	SMPL
J6548-06 - 1	D-6	452	0.1374	-7470A	PPB		27 Dec 2018 12:37:08	U	SMPL
J6548-07 - 1	D-7	436	0.1333	-7470A	PPB		27 Dec 2018 12:39:15	U	SMPL
J6548-08 - 1	D-8	430	0.1317	-7470A	PPB		27 Dec 2018 12:41:22	U	SMPL
J6548-09 - 1	D-9	413	0.1273	-7470A	PPB		27 Dec 2018 12:43:30	U	SMPL
J6548-10 - 1	D-10	454	0.1379	-7470A	PPB		27 Dec 2018 12:45:39	U	SMPL
J6548-11 - 1	D-11	423	0.1299	-7470A	PPB		27 Dec 2018 12:47:47	U	SMPL
J6548-12 - 1	D-12	422	0.1297	-7470A	PPB		27 Dec 2018 12:49:55	U	SMPL
CCV97 - 1	CCV97	18974	4.8997	-7470A	PPB		27 Dec 2018 12:52:03	U	SMPL
CCB97 - 1	CCB97	11	0.0240	-7470A	PPB		27 Dec 2018 12:54:10	U	SMPL
J6548-13 - 1	D-13	417	0.1284	-7470A	PPB		27 Dec 2018 12:56:18	U	SMPL
J6548-14 - 1	D-14	419	0.1289	-7470A	PPB		27 Dec 2018 12:58:25	U	SMPL
J6548-15 - 1	D-15	429	0.1315	-7470A	PPB		27 Dec 2018 13:00:32	U	SMPL
J6548-16 - 1	D-16	445	0.1356	-7470A	PPB		27 Dec 2018 13:02:40	U	SMPL
J6548-17 - 1	D-17	443	0.1351	-7470A	PPB		27 Dec 2018 13:04:47	U	SMPL
J6548-18 - 1	D-18	428	0.1312	-7470A	PPB		27 Dec 2018 13:06:54	U	SMPL
PB115987BL - 1	PBW	-10	0.0186	-7470A	PPB		27 Dec 2018 13:09:02	U	SMPL
PB115987BS - 1	LCSW	14436	3.7329	-7470A	PPB		27 Dec 2018 13:11:11	U	SMPL
J6533-07 - 1	COMP-1	-65	0.0044	-7470A	PPB		27 Dec 2018 13:13:20	U	SMPL
J6533-07DUP - 1	COMP-1DUP	-55	0.0070	-7470A	PPB		27 Dec 2018 13:15:28	U	SMPL
CCV98 - 1	CCV98	18839	4.8649	-7470A	PPB		27 Dec 2018 13:17:36	U	SMPL

LB100052 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Dec 2018 09:42:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB98 - 1	CCB98	-55	0.0070		-7470A	PPB	27 Dec 2018 13:19:44	U	SMPL
J6533-07MS - 1	COMP-1MS	15949	4.1219		-7470A	PPB	27 Dec 2018 13:21:52	U	SMPL
J6533-07MSD - 1	COMP-1MSD	14722	3.8064		-7470A	PPB	27 Dec 2018 13:23:59	U	SMPL
J6533-10 - 1	COMP-2	-88	-0.0015		-7470A	PPB	27 Dec 2018 13:26:07	U	SMPL
J6533-13 - 1	COMP-3	1348	0.3677		-7470A	PPB	27 Dec 2018 13:28:15	U	SMPL
J6533-16 - 1	COMP-4	90	0.0443		-7470A	PPB	27 Dec 2018 13:30:23	U	SMPL
J6533-19 - 1	COMP-5	26	0.0278		-7470A	PPB	27 Dec 2018 13:32:30	U	SMPL
PB115964TB - 1	PB115952TB	95	0.0456		-7470A	PPB	27 Dec 2018 13:37:10	U	SMPL
PB115965TB - 1	PB115955TB	135	0.0559		-7470A	PPB	27 Dec 2018 13:39:18	U	SMPL
PB115966TB - 1	PB115957TB	102	0.0474		-7470A	PPB	27 Dec 2018 13:41:27	U	SMPL
CCV99 - 1	CCV99	19000	4.9063		-7470A	PPB	27 Dec 2018 13:43:34	U	SMPL
CCB99 - 1	CCB99	85	0.0430		-7470A	PPB	27 Dec 2018 13:45:42	U	SMPL
J6188-06LX5 - 1		-7	0.0194		-7470A	PPB	27 Dec 2018 13:47:50	U	SMPL
J6188-06A - 1		15279	3.9496		-7470A	PPB	27 Dec 2018 13:49:58	U	SMPL
J6533-08LX5 - 1		140	0.0571		-7470A	PPB	27 Dec 2018 13:52:06	U	SMPL
J6533-08A - 1		14444	3.7349		-7470A	PPB	27 Dec 2018 13:54:14	U	SMPL
J6548-01LX5 - 1		186	0.0690		-7470A	PPB	27 Dec 2018 13:56:22	U	SMPL
J6548-01A - 1		14162	3.6624		-7470A	PPB	27 Dec 2018 13:58:30	U	SMPL
J6533-07LX5 - 1		159	0.0620		-7470A	PPB	27 Dec 2018 14:00:38	U	SMPL
J6533-07A - 1		16378	4.2322		-7470A	PPB	27 Dec 2018 14:02:46	U	SMPL
CCV01 - 1	CCV01	18871	4.8732		-7470A	PPB	27 Dec 2018 14:04:54	U	SMPL
CCB01 - 1	CCB01	13	0.0245		-7470A	PPB	27 Dec 2018 14:07:03	U	SMPL