



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB100056

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	12/26/2018	15:45
2	CAL2	1	Water			21.2	7.00	12/26/2018	15:47
3	CAL3	1	Water			21.3	10.04	12/26/2018	15:49
4	ICV	1	Water			21.1	7.01	12/26/2018	15:51
5	CCV1	1	Water			21.2	2.00	12/26/2018	16:30
6	J6195-12	1	Solid	20.02	20	20.3	7.67	12/26/2018	16:32
7	J6195-12DUP	1	Solid	20.04	20	20.4	7.68	12/26/2018	16:34
8	J6195-14	1	Solid	20.02	20	20.4	8.08	12/26/2018	16:36
9	J6579-03	1	Solid	20.03	20	20.1	5.35	12/26/2018	16:38
10	CCV2	1	Water			20.1	12.00	12/26/2018	16:40

WORKLIST(Hardcopy Internal Chain)

LB100856

WorkList Name : corrosivity122618

WorkList ID : 120783

Date : 12/26/2018 3:33:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2018	Solid	J6195-12	Corrosivity	Cool 4 deg C	PSEG05		OR-01-122618-A	12/26/2018	9045D
12/26/2018	Solid	J6195-14	Corrosivity	Cool 4 deg C	PSEG05		OR-01-122618-B	12/26/2018	9045D
12/26/2018	Solid	J6579-03	Corrosivity	Cool 4 deg C	PSEG01	E61	TN-122618-C	12/26/2018	9045D

Date/Time 12/26/18 4:15
 Received by: SP
 Relinquished by: CP

Date/Time 12/26/18 4:50
 Received by: CP
 Relinquished by: SP