

LB 1000 66


245.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37241



A= 0.0000e+000

B= 2.6912e-004

C= -2.9399e-002

Rho= 0.9997372

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	-9	0.000	-9				
0.05	0.050	0.037	-0.013	246	0.0 %	246				
0.2	0.200	0.184	-0.016	792	0.0 %	792				
2.5	2.500	2.505	0.005	9418	0.0 %	9418				
5.0	5.000	5.185	0.185	19376	0.0 %	19376				
7.5	7.500	7.378	-0.122	27524	0.0 %	27524				
10.0	10.000	9.993	-0.007	37241	0.0 %	37241				

LB100066 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Dec 2018 14:37:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.0000	245.1	PPB	27 Dec 2018 14:41:10	S	Std
0.05 - 1	5005	246	-	0.0500	245.1	PPB	27 Dec 2018 14:43:28	S	Std
0.2 - 1	50.2	792	-	0.2000	245.1	PPB	27 Dec 2018 14:45:46	S	Std
2.5 - 1	52.5	9418	-	2.5000	245.1	PPB	27 Dec 2018 14:48:03	S	Std
5.0 - 1	55	19376	-	5.0000	245.1	PPB	27 Dec 2018 14:50:20	S	Std
7.5 - 1	57.5	27524	-	7.5000	245.1	PPB	27 Dec 2018 14:52:37	S	Std
10.0 - 1	510	37241	-	10.0000	245.1	PPB	27 Dec 2018 15:00:10	S	Std
ICV43 - 1	ICV43	14977	4.0013	-245.1		PPB	27 Dec 2018 15:05:36	U	SMPL
ICB43 - 1	ICB43	-135	-0.0657	-245.1		PPB	27 Dec 2018 15:07:51	U	SMPL
CCV02 - 1	CCV02	19075	5.1041	-245.1		PPB	27 Dec 2018 15:10:07	U	SMPL
CCB02 - 1	CCB02	-141	-0.0673	-245.1		PPB	27 Dec 2018 15:12:23	U	SMPL
CRA - 1	CRA	252	0.0384	-245.1		PPB	27 Dec 2018 15:14:39	U	SMPL
HighStd - 1	HighStd	37236	9.9917	-245.1		PPB	27 Dec 2018 15:16:55	U	SMPL
ChkStd - 1	ChkStd	25536	6.8429	-245.1		PPB	27 Dec 2018 15:19:11	U	SMPL
PB116030BL - 1	PBW	-51	-0.0431	-245.1		PPB	27 Dec 2018 15:24:42	U	SMPL
PB116030BS - 1	LCSW	13736	3.6673	-245.1		PPB	27 Dec 2018 15:26:58	U	SMPL
J6502-01 - 1	121918-JETT-E-D-B	-81	-0.0512	-245.1		PPB	27 Dec 2018 15:29:15	U	SMPL
J6502-02 - 1	121918-JETT-E-D-S	-37	-0.0394	-245.1		PPB	27 Dec 2018 15:31:32	U	SMPL
J6502-03 - 1	121918-JETT-E-U-B	-57	-0.0447	-245.1		PPB	27 Dec 2018 15:33:50	U	SMPL
J6502-04 - 1	121918-JETT-E-U-S	-58	-0.0450	-245.1		PPB	27 Dec 2018 15:36:07	U	SMPL
J6502-04DUP - 1	121918-JETT-E-U-SDUP	-64	-0.0466	-245.1		PPB	27 Dec 2018 15:38:23	U	SMPL
CCV03 - 1	CCV03	19090	5.1082	-245.1		PPB	27 Dec 2018 15:40:40	U	SMPL
CCB03 - 1	CCB03	-127	-0.0636	-245.1		PPB	27 Dec 2018 15:42:55	U	SMPL
J6502-04MS - 1	121918-JETT-E-U-SMS	15210	4.0640	-245.1		PPB	27 Dec 2018 15:45:11	U	SMPL
J6502-04MSD - 1	121918-JETT-E-U-SMSD	14802	3.9542	-245.1		PPB	27 Dec 2018 15:47:27	U	SMPL
J6502-04LX5 - 1		-119	-0.0614	-245.1		PPB	27 Dec 2018 15:49:43	U	SMPL
J6502-04A - 1		16977	4.5395	-245.1		PPB	27 Dec 2018 15:52:00	U	SMPL
CCV04 - 1	CCV04	18895	5.0557	-245.1		PPB	27 Dec 2018 15:54:16	U	SMPL
CCB04 - 1	CCB04	-110	-0.0590	-245.1		PPB	27 Dec 2018 15:56:33	U	SMPL