

LB 100068


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52580



A= 0.0000e+000

B= 1.9096e-004

C= -4.1963e-002

Rho= 0.9998996

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.050	-0.050	-41	0.000	-41				
0.2	0.200	0.176	-0.024	1139	0.0 %	1139				
2.5	2.500	2.599	0.099	13827	0.0 %	13827				
5.0	5.000	5.029	0.029	26556	0.0 %	26556				
7.5	7.500	7.447	-0.053	39219	0.0 %	39219				
10.0	10.000	9.999	-0.001	52580	0.0 %	52580				

LB100068 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Dec 2018 16:43:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7471B	PPB	27 Dec 2018 16:47:53	S	Std
0.2 - 1	50.2	1139	-	0.2000	7471B	PPB	27 Dec 2018 16:50:11	S	Std
2.5 - 1	52.5	13827	-	2.5000	7471B	PPB	27 Dec 2018 16:52:29	S	Std
5.0 - 1	55	26556	-	5.0000	7471B	PPB	27 Dec 2018 16:54:48	S	Std
7.5 - 1	57.5	39219	-	7.5000	7471B	PPB	27 Dec 2018 16:57:05	S	Std
10.0 - 1	510	52580	-	10.0000	7471B	PPB	27 Dec 2018 16:59:23	S	Std
ICV44 - 1	ICV44	20678	3.9068	-	7471B	PPB	27 Dec 2018 17:02:08	U	SMPL
ICB44 - 1	ICB44	-177	-0.0758	-	7471B	PPB	27 Dec 2018 17:04:23	U	SMPL
CCV05 - 1	CCV05	26402	4.9999	-	7471B	PPB	27 Dec 2018 17:06:40	U	SMPL
CCB05 - 1	CCB05	-176	-0.0756	-	7471B	PPB	27 Dec 2018 17:08:56	U	SMPL
CRA - 1	CRA	1112	0.1704	-	7471B	PPB	27 Dec 2018 17:11:13	U	SMPL
HighStd - 1	HighStd	52754	10.0322	-	7471B	PPB	27 Dec 2018 17:13:29	U	SMPL
ChkStd - 1	ChkStd	36356	6.9008	-	7471B	PPB	27 Dec 2018 17:15:45	U	SMPL
PB116002BL - 1	PBS	-195	-0.0792	-	7471B	PPB	27 Dec 2018 17:18:04	U	SMPL
PB116002BS - 1	LCSS	19054	3.5967	-	7471B	PPB	27 Dec 2018 17:20:22	U	SMPL
J6195-11 - 1	OR-01-122618-A	1503	0.2451	-	7471B	PPB	27 Dec 2018 17:22:40	U	SMPL
J6195-13 - 1	OR-01-122618-B	1768	0.2957	-	7471B	PPB	27 Dec 2018 17:24:57	U	SMPL
J6547-02 - 1	COMP-1A	12005	2.2506	-	7471B	PPB	27 Dec 2018 17:27:15	U	SMPL
J6547-04 - 1	COMP-1B	7521	1.3943	-	7471B	PPB	27 Dec 2018 17:29:34	U	SMPL
J6547-04DUP - 1	COMP-1BDUP	8119	1.5085	-	7471B	PPB	27 Dec 2018 17:31:49	U	SMPL
CCV06 - 1	CCV06	26214	4.9640	-	7471B	PPB	27 Dec 2018 17:34:05	U	SMPL
CCB06 - 1	CCB06	-192	-0.0786	-	7471B	PPB	27 Dec 2018 17:36:21	U	SMPL
J6547-04MS - 1	COMP-1BMS	29554	5.6018	-	7471B	PPB	27 Dec 2018 17:38:38	U	SMPL
J6547-04MSD - 1	COMP-1BMSD	27973	5.2999	-	7471B	PPB	27 Dec 2018 17:40:54	U	SMPL
J6579-01 - 1	TN-122618-C	1415	0.2283	-	7471B	PPB	27 Dec 2018 17:43:11	U	SMPL
J6547-04LX5 - 1		1312	0.2086	-	7471B	PPB	27 Dec 2018 17:45:29	U	SMPL
J6547-04A - 1		29194	5.5331	-	7471B	PPB	27 Dec 2018 17:47:46	U	SMPL
CCV07 - 1	CCV07	26472	5.0133	-	7471B	PPB	27 Dec 2018 17:50:03	U	SMPL
CCB07 - 1	CCB07	-192	-0.0786	-	7471B	PPB	27 Dec 2018 17:52:22	U	SMPL