

LB 1000 76


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

39305



A= 0.0000e+000

B= 2.5273e-004

C= 6.4566e-003

Rho= 0.9998312

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	-31	0.000	-31				
0.2	0.200	0.199	-0.001	762	0.0 %	762				
2.5	2.500	2.415	-0.085	9530	0.0 %	9530				
5.0	5.000	5.128	0.128	20265	0.0 %	20265				
7.5	7.500	7.519	0.019	29725	0.0 %	29725				
10.0	10.000	9.940	-0.060	39305	0.0 %	39305	0			

LB100076 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Dec 2018 09:07:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-31	-	0.0000	7470A	PPB	28 Dec 2018 09:18:52	S	Std
0.2 - 1	50.2	762	-	0.2000	7470A	PPB	28 Dec 2018 09:21:01	S	Std
2.5 - 1	52.5	9530	-	2.5000	7470A	PPB	28 Dec 2018 09:23:10	S	Std
5.0 - 1	55	20265	-	5.0000	7470A	PPB	28 Dec 2018 09:25:18	S	Std
7.5 - 1	57.5	29725	-	7.5000	7470A	PPB	28 Dec 2018 09:27:26	S	Std
10.0 - 1	510	39305	-	10.0000	7470A	PPB	28 Dec 2018 09:32:06	S	Std
ICV45 - 1	ICV45	14831	3.7548	-	7470A	PPB	28 Dec 2018 09:34:41	U	SMPL
ICB45 - 1	ICB45	-126	-0.0254	-	7470A	PPB	28 Dec 2018 09:36:47	U	SMPL
CCV08 - 1	CCV08	19597	4.9593	-	7470A	PPB	28 Dec 2018 09:38:54	U	SMPL
CCB08 - 1	CCB08	-113	-0.0221	-	7470A	PPB	28 Dec 2018 09:41:00	U	SMPL
CRA - 1	CRA	774	0.2021	-	7470A	PPB	28 Dec 2018 09:43:08	U	SMPL
HighStd - 1	HighStd	38686	9.7838	-	7470A	PPB	28 Dec 2018 09:45:15	U	SMPL
ChkStd - 1	ChkStd	28039	7.0929	-	7470A	PPB	28 Dec 2018 09:47:22	U	SMPL
PB116037BL - 1	PBW	-38	-0.0031	-	7470A	PPB	28 Dec 2018 09:59:48	U	SMPL
PB116037BS - 1	LCSW	14647	3.7083	-	7470A	PPB	28 Dec 2018 10:02:25	U	SMPL
J6195-12 - 1	OR-01-122618-A	-85	-0.0150	-	7470A	PPB	28 Dec 2018 10:04:32	U	SMPL
J6195-12DUP - 1	OR-01-122618-ADUP	-1	0.0062	-	7470A	PPB	28 Dec 2018 10:06:40	U	SMPL
J6195-12MS - 1	OR-01-122618-AMS	14001	3.5450	-	7470A	PPB	28 Dec 2018 10:08:48	U	SMPL
J6195-12MSD - 1	OR-01-122618-AMSD	13514	3.4219	-	7470A	PPB	28 Dec 2018 10:10:57	U	SMPL
J6195-14 - 1	OR-01-122618-B	-106	-0.0203	-	7470A	PPB	28 Dec 2018 10:13:04	U	SMPL
CCV09 - 1	CCV09	19525	4.9411	-	7470A	PPB	28 Dec 2018 10:15:10	U	SMPL
CCB09 - 1	CCB09	-104	-0.0198	-	7470A	PPB	28 Dec 2018 10:17:17	U	SMPL
J6571-01 - 1	50072	-34	-0.0021	-	7470A	PPB	28 Dec 2018 10:19:24	U	SMPL
J6571-02 - 1	10827	-32	-0.0016	-	7470A	PPB	28 Dec 2018 10:21:31	U	SMPL
J6579-03 - 1	TN-122618-C	26	0.0130	-	7470A	PPB	28 Dec 2018 10:23:38	U	SMPL
PB116038BL - 1	PBW	-24	0.0004	-	7470A	PPB	28 Dec 2018 10:25:45	U	SMPL
PB116038BS - 1	LCSW	14781	3.7421	-	7470A	PPB	28 Dec 2018 10:27:53	U	SMPL
J6571-03 - 1	30061	-110	-0.0213	-	7470A	PPB	28 Dec 2018 10:30:01	U	SMPL
J6571-04 - 1	50082-50083	87	0.0284	-	7470A	PPB	28 Dec 2018 10:32:08	U	SMPL
J6571-05 - 1	50084-50085-50086	-62	-0.0092	-	7470A	PPB	28 Dec 2018 10:34:17	U	SMPL
J6571-09 - 1	50087	-10	0.0039	-	7470A	PPB	28 Dec 2018 10:36:25	U	SMPL
J6575-02 - 1	SS-122618-A	-55	-0.0074	-	7470A	PPB	28 Dec 2018 10:38:32	U	SMPL
CCV10 - 1	CCV10	19604	4.9611	-	7470A	PPB	28 Dec 2018 10:40:39	U	SMPL
CCB10 - 1	CCB10	-154	-0.0325	-	7470A	PPB	28 Dec 2018 10:42:46	U	SMPL
J6575-02DUP - 1	SS-122618-ADUP	-93	-0.0170	-	7470A	PPB	28 Dec 2018 10:44:53	U	SMPL
J6575-02MS - 1	SS-122618-AMS	14679	3.7163	-	7470A	PPB	28 Dec 2018 10:50:03	U	SMPL
J6575-02MSD - 1	SS-122618-AMSD	14747	3.7335	-	7470A	PPB	28 Dec 2018 10:52:10	U	SMPL
PB116039BL - 1	PBW	-118	-0.0234	-	7470A	PPB	28 Dec 2018 10:54:17	U	SMPL
PB116039BS - 1	LCSW	14562	3.6868	-	7470A	PPB	28 Dec 2018 10:56:24	U	SMPL
J6549-01 - 1	TW-1	1964	0.5028	-	7470A	PPB	28 Dec 2018 10:58:32	U	SMPL
J6549-01DUP - 1	TW-1DUP	1724	0.4422	-	7470A	PPB	28 Dec 2018 11:00:41	U	SMPL
J6549-01MS - 1	TW-1MS	17219	4.3583	-	7470A	PPB	28 Dec 2018 11:02:49	U	SMPL
J6549-01MSD - 1	TW-1MSD	17821	4.5104	-	7470A	PPB	28 Dec 2018 11:07:27	U	SMPL
J6549-02 - 1	MW-6	-68	-0.0107	-	7470A	PPB	28 Dec 2018 11:09:34	U	SMPL
CCV11 - 1	CCV11	19553	4.9482	-	7470A	PPB	28 Dec 2018 11:11:41	U	SMPL
CCB11 - 1	CCB11	-127	-0.0256	-	7470A	PPB	28 Dec 2018 11:13:48	U	SMPL
J6549-03 - 1	MW-7	207	0.0588	-	7470A	PPB	28 Dec 2018 11:15:56	U	SMPL
J6549-04 - 1	MW-8	56	0.0206	-	7470A	PPB	28 Dec 2018 11:18:03	U	SMPL
J6549-05 - 1	MW-9	185	0.0532	-	7470A	PPB	28 Dec 2018 11:20:10	U	SMPL
J6549-06 - 1	MW-11	-66	-0.0102	-	7470A	PPB	28 Dec 2018 11:22:18	U	SMPL
J6549-07 - 1	FIELD-BLANK	-82	-0.0143	-	7470A	PPB	28 Dec 2018 11:24:25	U	SMPL
J6549-08 - 1	DUP-1	-24	0.0004	-	7470A	PPB	28 Dec 2018 11:26:33	U	SMPL
J6563-01 - 1	P001-RB-181218	-92	-0.0168	-	7470A	PPB	28 Dec 2018 11:28:41	U	SMPL
J6563-02 - 1	P001-RB-181219	-29	-0.0009	-	7470A	PPB	28 Dec 2018 11:30:50	U	SMPL
J6563-03 - 1	P001-RB-181220	-33	-0.0019	-	7470A	PPB	28 Dec 2018 11:32:58	U	SMPL
J6566-14 - 1	P001-RB-181217	-99	-0.0186	-	7470A	PPB	28 Dec 2018 11:35:06	U	SMPL
CCV12 - 1	CCV12	19598	4.9596	-	7470A	PPB	28 Dec 2018 11:37:13	U	SMPL
CCB12 - 1	CCB12	-148	-0.0309	-	7470A	PPB	28 Dec 2018 11:39:20	U	SMPL
J6575-01 - 1	SS-122618-A	142	0.0423	-	7470A	PPB	28 Dec 2018 11:41:27	U	SMPL
PB116037TB - 1	PB115998TB	-87	-0.0155	-	7470A	PPB	28 Dec 2018 11:43:34	U	SMPL
PB116038TB - 1	PB115001TB	-18	0.0019	-	7470A	PPB	28 Dec 2018 11:45:41	U	SMPL
J6195-12LX5 - 1		14	0.0100	-	7470A	PPB	28 Dec 2018 11:47:48	U	SMPL
J6195-12A - 1		15115	3.8265	-	7470A	PPB	28 Dec 2018 11:49:56	U	SMPL
J6575-02LX5 - 1		-102	-0.0193	-	7470A	PPB	28 Dec 2018 11:52:03	U	SMPL
J6575-02A - 1		13621	3.4490	-	7470A	PPB	28 Dec 2018 11:59:16	U	SMPL
J6549-01LX5 - 1		262	0.0727	-	7470A	PPB	28 Dec 2018 12:01:24	U	SMPL
J6549-01A - 1		17967	4.5473	-	7470A	PPB	28 Dec 2018 12:03:32	U	SMPL
CCV13 - 1	CCV13	19253	4.8724	-	7470A	PPB	28 Dec 2018 12:05:40	U	SMPL
CCB13 - 1	CCB13	-117	-0.0231	-	7470A	PPB	28 Dec 2018 12:07:48	U	SMPL