



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100077

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.0       | 80.6      | 81.2            | 12/28/2018 | 08:45     |
| 2   | J6533-23    |               | 1  | 2                 | 22.0       | 71.6      | 72.2            | 12/28/2018 | 09:01     |
| 3   | J6533-23DUP |               | 1  | 2                 | 22.0       | 71.6      | 72.2            | 12/28/2018 | 09:11     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

CB 100077

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP122718-2

WorkList ID : 120830

Date : 12/27/2018 11:34:42 AM

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 12/26/2018 | Solid  | J6533-23 | Flash Point | Cool 4 deg C | PSEG01   | E52              | 09017           | 12/21/2018   | 1010A  |

Date/Time 12-28-18 8:30  
 Received by: AK  
 Relinquished by: AK

Date/Time 12-28-18 9:12  
 Received by: CP  
 Relinquished by: AK