

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 9:03

Test: Total CN

Sample Id	Result <i>mg/d</i>	Dil. 1 +	Response	Errors
ICV1	97.547	0.0	0.091	
ICB1	-0.105	0.0	0.004	
CCV1	242.376	0.0	0.220	
CCB1	-0.351	0.0	0.004	
PB116019BL	-0.334	0.0	0.004	
PB116019BS	98.620	0.0	0.092	
LOWPB116019	9.538	0.0	0.013	
HIGHPB116019	487.767	0.0	0.439	
J6195-11MS	32.923	0.0	0.034	
J6195-11MSD	32.287	0.0	0.033	
J6195-13	1.397	0.0	0.006	
J6579-01	1.126	0.0	0.005	
CCV2	241.507	0.0	0.220	
CCB2	-0.316	0.0	0.004	
PB116020BL	-0.428	0.0	0.004	
PB116020BS	98.809	0.0	0.092	
J6575-01	0.890	0.0	0.005	
J6575-01DUP	0.865	0.0	0.005	
J6575-01MS	35.603	0.0	0.036	
J6575-01MSD	36.147	0.0	0.037	
J6195-11	-0.466	0.0	0.004	
J6195-11DUP	-0.214	0.0	0.004	
CCV3	255.094	0.0	0.232	
CCB3	-0.220	0.0	0.004	

N	24
Mean	69.586
SD	121.1629
CV%	174.12

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 8:02


Test Total CN

Accepted 12/28/2018 8:02

Factor 1122
Bias 0.004

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9065	0.0000	
2	5.0PPBCN	0.008	4.1588	5.0000	
3	10PPBCN	0.012	9.0998	10.0000	
4	50PPBCN	0.049	49.9548	50.0000	
5	100PPBCN	0.095	102.2571	100.0000	
6	250PPBCN	0.229	251.7130	250.0000	
7	500PPBCN	0.449	498.7231	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 28 07:44:26 2018

Fri Dec 28 09:01:58 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9065	µg/l	12/28/2018 8:01:14	
5.0PPBCN	A	Total CN	P	4.1588	µg/l	12/28/2018 8:01:15	
10PPBCN	A	Total CN	P	9.0998	µg/l	12/28/2018 8:01:16	
50PPBCN	A	Total CN	P	49.9548	µg/l	12/28/2018 8:01:17	
100PPBCN	A	Total CN	P	102.2571	µg/l	12/28/2018 8:01:18	
250PPBCN	A	Total CN	P	251.713	µg/l	12/28/2018 8:01:19	
500PPBCN	A	Total CN	P	498.7231	µg/l	12/28/2018 8:01:20	
ICV1	S	Total CN	P	97.5469	µg/l	12/28/2018 8:31:42	
ICB1	S	Total CN	P	-0.1045	µg/l	12/28/2018 8:31:43	
CCV1	S	Total CN	P	242.376	µg/l	12/28/2018 8:31:44	
CCB1	S	Total CN	P	-0.3508	µg/l	12/28/2018 8:31:45	
PB116019BL	S	Total CN	P	-0.3341	µg/l	12/28/2018 8:31:46	
PB116019BS	S	Total CN	P	98.6202	µg/l	12/28/2018 8:31:47	
LOWPB116019	S	Total CN	P	9.5382	µg/l	12/28/2018 8:31:48	
HIGHPB116019	S	Total CN	P	487.7672	µg/l	12/28/2018 8:31:49	
J6195-11MS	S	Total CN	P	32.9234	µg/l	12/28/2018 8:31:52	
J6195-11MSD	S	Total CN	P	32.2871	µg/l	12/28/2018 8:39:17	
J6195-13	S	Total CN	P	1.3967	µg/l	12/28/2018 8:39:18	
J6579-01	S	Total CN	P	1.1257	µg/l	12/28/2018 8:39:19	
CCV2	S	Total CN	P	241.5071	µg/l	12/28/2018 8:39:20	
CCB2	S	Total CN	P	-0.316	µg/l	12/28/2018 8:39:21	
PB116020BL	S	Total CN	P	-0.4282	µg/l	12/28/2018 8:39:22	
PB116020BS	S	Total CN	P	98.8092	µg/l	12/28/2018 8:39:23	
J6575-01	S	Total CN	P	0.8903	µg/l	12/28/2018 8:39:24	
J6575-01DUP	S	Total CN	P	0.865	µg/l	12/28/2018 8:39:25	
J6575-01MS	S	Total CN	P	35.6034	µg/l	12/28/2018 8:45:28	
J6575-01MSD	S	Total CN	P	36.1468	µg/l	12/28/2018 8:45:29	
J6195-11	S	Total CN	P	-0.4663	µg/l	12/28/2018 9:01:55	
J6195-11DUP	S	Total CN	P	-0.2136	µg/l	12/28/2018 9:01:56	
CCV3	S	Total CN	P	255.0936	µg/l	12/28/2018 9:01:57	
CCB3	S	Total CN	P	-0.2204	µg/l	12/28/2018 9:01:58	