

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/27/2018

TEMP1 IN: 103 12/27/2018 13:00 TEMP OUT: 104 12/27/2018 14:10

TEMP2 IN: 104 12/27/2018 15:00 TEMP OUT: 104 12/28/2018 07:00

TEMP3 IN: 105 12/28/2018 07:40 TEMP OUT: 105 12/28/2018 09:15

TEMP4 IN: 105 12/28/2018 10:00 TEMP OUT: 105 12/28/2018 11:35

Run Number: LB100079

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB100079BL	LB100079BL	1.3954	1.3954	100	1.3954	1.3954	1.3954	0.0000	0
2	LB100079BS	LB100079BS	1.4028	1.4028	100	1.4533	1.4533	1.4533	0.0505	505
3	J6582-01	SPDES-1	1.4088	1.4088	950	1.4092	1.4092	1.4092	0.0004	0.4

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS122818

WorkList ID : 120842

Date : 12/28/2018 7:36:56 AM

12/28/18 09:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/28/2018	Water	J6582-01	TSS	Cool 4 deg C	SKAN06	D21	SPDES-1	12/27/2018	SM2540 D

FP-12791048

Date/Time 12/28/18, 07:40
 Received by: AS
 Relinquished by: [Signature]

Date/Time 12/28/18, 09:00
 Received by: [Signature]
 Relinquished by: AS