



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : sweta

Run Number: LB100081

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	12/28/2018	10:15
2	CAL2	1	Water			21.3	7.00	12/28/2018	10:16
3	CAL3	1	Water			21.3	10.03	12/28/2018	10:17
4	ICV	1	Water			21.3	6.99	12/28/2018	10:18
5	CCV1	1	Water			21.2	2.01	12/28/2018	10:20
6	J6193-12	1	Solid	20.02	20	21.1	8.14	12/28/2018	10:21
7	J6193-12DUP	1	Solid	20.03	20	21.2	8.15	12/28/2018	10:22
8	J6193-14	1	Solid	20.04	20	20.8	8.40	12/28/2018	10:23
9	J6193-16	1	Solid	20.01	20	20.4	8.31	12/28/2018	10:25
10	J6193-18	1	Solid	20.03	20	20.1	8.22	12/28/2018	10:27
11	J6584-01	1	Solid	20.01	20	19.6	7.22	12/28/2018	10:30
12	J6584-02	1	Solid	20.03	20	19.2	6.71	12/28/2018	10:32
13	J6584-03	1	Solid	20.04	20	8.30	19.5	12/28/2018	10:33
14	CCV2	1	Water			21.3	12.01	12/28/2018	10:35

LB (0008)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j6584

WorkList ID : 120867

Date : 12/28/2018 10:09:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/27/2018	Solid	J6193-12	Corrosivity	Cool 4 deg C	PSEG05		CL-02-122718-A	12/27/2018	9045D
12/27/2018	Solid	J6193-14	Corrosivity	Cool 4 deg C	PSEG05		CL-02-122718-B	12/27/2018	9045D
12/27/2018	Solid	J6193-16	Corrosivity	Cool 4 deg C	PSEG05		CL-02-122718-C	12/27/2018	9045D
12/27/2018	Solid	J6193-18	Corrosivity	Cool 4 deg C	PSEG05		CL-02-122718-D	12/27/2018	9045D
12/26/2018	Solid	J6584-01	Corrosivity	Cool 4 deg C	CLEA11	D22	ENGINE-WASH	12/26/2018	9045D
12/26/2018	Solid	J6584-02	Corrosivity	Cool 4 deg C	CLEA11	D22	RECLAIMER-PIT	12/26/2018	9045D
12/26/2018	Solid	J6584-03	Corrosivity	Cool 4 deg C	CLEA11	D22	BUS-WASH	12/26/2018	9045D

Date/Time 12-28-18 10:10 AM

Received by: Ju

Relinquished by: CB

Date/Time 12-28-18 11:30 AM

Received by: CB

Relinquished by: Ju