

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 11:09

Reactive
Test: ~~Total~~ CN

Sample Id	<i>Mg/d</i> Result	Dil. 1 +	Response	Errors
ICV1	96.839	0.0	0.091	
ICB1	-0.356	0.0	0.004	
CCV1	241.161	0.0	0.219	
CCB1	-0.499	0.0	0.004	
PB116062BL	-0.462	0.0	0.004	
J6193-12	-0.668	0.0	0.004	
J6193-12DUP	-0.523	0.0	0.004	
J6193-14	-0.552	0.0	0.004	
J6193-16	-0.464	0.0	0.004	
J6193-18	-0.559	0.0	0.004	
CCV2	239.991	0.0	0.218	
CCB2	-0.892	0.0	0.004	

N	12
Mean	47.751
SD	94.2786
CV%	197.44

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 28 11:06:03 2018

Fri Dec 28 11:08:51 2018

A7

Sample Id	Sam/Ctr/c/	Test short	nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive	CN	P	-0.9065	µg/l	12/28/2018 8:01:14	
5.OPPBCN	A	Reactive	CN	P	4.1588	µg/l	12/28/2018 8:01:15	
10PPBCN	A	Reactive	CN	P	9.0998	µg/l	12/28/2018 8:01:16	
50PPBCN	A	Reactive	CN	P	49.9548	µg/l	12/28/2018 8:01:17	
100PPBCN	A	Reactive	CN	P	102.2571	µg/l	12/28/2018 8:01:18	
250PPBCN	A	Reactive	CN	P	251.713	µg/l	12/28/2018 8:01:19	
500PPBCN	A	Reactive	CN	P	498.7231	µg/l	12/28/2018 8:01:20	
ICV1	S	Reactive	CN	P	96.8394	µg/l	12/28/2018 11:06:03	
ICB1	S	Reactive	CN	P	-0.3557	µg/l	12/28/2018 11:06:04	
CCV1	S	Reactive	CN	P	241.1615	µg/l	12/28/2018 11:06:05	
CCB1	S	Reactive	CN	P	-0.499	µg/l	12/28/2018 11:06:06	
PB116062BL	S	Reactive	CN	P	-0.4622	µg/l	12/28/2018 11:06:07	
J6193-12	S	Reactive	CN	P	-0.6682	µg/l	12/28/2018 11:06:08	
J6193-12DUP	S	Reactive	CN	P	-0.5233	µg/l	12/28/2018 11:06:09	
J6193-14	S	Reactive	CN	P	-0.552	µg/l	12/28/2018 11:06:10	
J6193-16	S	Reactive	CN	P	-0.4643	µg/l	12/28/2018 11:06:11	
J6193-18	S	Reactive	CN	P	-0.5592	µg/l	12/28/2018 11:06:12	
CCV2	S	Reactive	CN	P	239.9908	µg/l	12/28/2018 11:06:13	
CCB2	S	Reactive	CN	P	-0.892	µg/l	12/28/2018 11:08:51	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

12/28/2018 11:10


Test Total CN

Accepted 12/28/2018 8:02

Factor 1122
 Bias 0.004

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9065	0.0000	
2	5.0PPBCN	0.008	4.1588	5.0000	
3	10PPBCN	0.012	9.0998	10.0000	
4	50PPBCN	0.049	49.9548	50.0000	
5	100PPBCN	0.095	102.2571	100.0000	
6	250PPBCN	0.229	251.7130	250.0000	
7	500PPBCN	0.449	498.7231	500.0000	