

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 11:52

Test: Total CN

Sample Id	Result ^{ug/l}	Dil. 1 +	Response	Errors
ICV1	96.860	0.0	0.091	
ICB1	-0.165	0.0	0.004	
CCV1	242.204	0.0	0.220	
CCB1	-0.511	0.0	0.004	
PB116064BL	-0.360	0.0	0.004	
PB116064BS	97.236	0.0	0.091	
LOWPB116064	9.475	0.0	0.013	
HIGHPB116064	484.981	0.0	0.437	
J6193-11	1.637	0.0	0.006	
J6193-11DUP	1.305	0.0	0.006	
J6193-11MS	29.400	0.0	0.031	
J6193-11MSD	28.267	0.0	0.030	
J6193-13	4.205	0.0	0.008	
J6193-15	1.953	0.0	0.006	
CCV2	239.033	0.0	0.217	
CCB2	-0.280	0.0	0.004	
J6193-17	1.901	0.0	0.006	
CCV3	243.334	0.0	0.221	
CCB3	-0.280	0.0	0.004	
N	19			
Mean	77.905			
SD	132.3973			
CV%	169.95			

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 11:10


Test Total CN

Accepted 12/28/2018 8:02

Factor 1122
Bias 0.004

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9065	0.0000	
2	5.0PPBCN	0.008	4.1588	5.0000	
3	10PPBCN	0.012	9.0998	10.0000	
4	50PPBCN	0.049	49.9548	50.0000	
5	100PPBCN	0.095	102.2571	100.0000	
6	250PPBCN	0.229	251.7130	250.0000	
7	500PPBCN	0.449	498.7231	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 28 11:39:32 2018

Fri Dec 28 11:49:55 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9065	µg/l	12/28/2018 8:01:14	
5.0PPBCN	A	Total CN	P	4.1588	µg/l	12/28/2018 8:01:15	
10PPBCN	A	Total CN	P	9.0998	µg/l	12/28/2018 8:01:16	
50PPBCN	A	Total CN	P	49.9548	µg/l	12/28/2018 8:01:17	
100PPBCN	A	Total CN	P	102.2571	µg/l	12/28/2018 8:01:18	
250PPBCN	A	Total CN	P	251.713	µg/l	12/28/2018 8:01:19	
500PPBCN	A	Total CN	P	498.7231	µg/l	12/28/2018 8:01:20	
ICV1	S	Total CN	P	96.8604	µg/l	12/28/2018 11:39:32	
ICB1	S	Total CN	P	-0.1651	µg/l	12/28/2018 11:39:33	
CCV1	S	Total CN	P	242.2036	µg/l	12/28/2018 11:39:34	
CCB1	S	Total CN	P	-0.5108	µg/l	12/28/2018 11:39:35	
PB116064BL	S	Total CN	P	-0.3597	µg/l	12/28/2018 11:39:36	
PB116064BS	S	Total CN	P	97.236	µg/l	12/28/2018 11:39:37	
LOWPB116064	S	Total CN	P	9.4752	µg/l	12/28/2018 11:39:38	
HIGHPB116064	S	Total CN	P	484.9814	µg/l	12/28/2018 11:39:39	
J6193-11	S	Total CN	P	1.6374	µg/l	12/28/2018 11:39:40	
J6193-11DUP	S	Total CN	P	1.3054	µg/l	12/28/2018 11:39:41	
J6193-11MS	S	Total CN	P	29.4003	µg/l	12/28/2018 11:39:42	
J6193-11MSD	S	Total CN	P	28.2668	µg/l	12/28/2018 11:47:07	
J6193-13	S	Total CN	P	4.2052	µg/l	12/28/2018 11:47:08	
J6193-15	S	Total CN	P	1.9529	µg/l	12/28/2018 11:47:09	
CCV2	S	Total CN	P	239.033	µg/l	12/28/2018 11:47:10	
CCB2	S	Total CN	P	-0.2796	µg/l	12/28/2018 11:47:11	
J6193-17	S	Total CN	P	1.9007	µg/l	12/28/2018 11:47:12	
CCV3	S	Total CN	P	243.3339	µg/l	12/28/2018 11:47:15	
CCB3	S	Total CN	P	-0.2797	µg/l	12/28/2018 11:47:16	