



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100094

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	12/28/2018	16:30
2	CAL2	1	Water			21.1	7.01	12/28/2018	16:31
3	CAL3	1	Water			21.1	10.04	12/28/2018	16:32
4	ICV	1	Water			21.1	6.99	12/28/2018	16:34
5	CCV1	1	Water			21.1	2.02	12/28/2018	16:35
6	J6595-01	1	Solid	20.03	20	20.2	9.38	12/28/2018	16:36
7	J6595-01DUP	1	Solid	20.04	20	20.3	9.39	12/28/2018	16:38
8	CCV2	1	Water			21.1	12.01	12/28/2018	16:39

1B100094

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j6595

WorkList ID : 120880

Date : 12/28/2018 4:23:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2018	Solid	J6595-01	Corrosivity	Cool 4 deg C	DIST01	D41	NR38-C1-110718	12/07/2018	9045D

Date/Time 12.28.18 4:25 PM
Received by: JP
Relinquished by: LG

Date/Time 12.28.18 5:23 PM
Received by: CF
Relinquished by: 30