



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100097

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	12/28/2018	16:30
2	CAL2	1	Water			21.1	7.01	12/28/2018	16:31
3	CAL3	1	Water			21.1	10.04	12/28/2018	16:32
4	ICV	1	Water			21.1	6.99	12/28/2018	16:34
5	CCV1	1	Water			21.2	2.01	12/28/2018	16:45
6	J6587-01	1	Water			17.8	6.90	12/28/2018	16:46
7	J6587-01DUP	1	Water			17.9	6.91	12/28/2018	16:47
8	CCV2	1	Water			21.1	12.02	12/28/2018	16:50

LB 100097

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivityW J6587

WorkList ID : 120881

Date : 12/28/2018 4:24:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/27/2018	Water	J6587-01	Corrosivity	Cool 4 deg C	PARS02		WC-20181227	12/27/2018	9040C

Date/Time 12-28-18 4:30 PM
Received by: JD
Relinquished by: JD

Date/Time 12-28-18 5:30 PM
Received by: JD
Relinquished by: JD