

LB1001017


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4620E



A= 0.0000e+000

B= 2.1673e-004

C= -2.9397e-002

Rho= 0.9999794

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	14	0.000	14				
0.2	0.200	0.187	-0.013	999	0.0 %	999				
2.5	2.500	2.538	0.038	11846	0.0 %	11846				
5.0	5.000	5.027	0.027	23330	0.0 %	23330				
7.5	7.500	7.490	-0.010	34693	0.0 %	34693				
10.0	10.000	9.985	-0.015	46206	0.0 %	46206				

LB1001017 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Feb 2019 09:25:25

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	14	-	0.00007470A	PPB	26 Feb 2019 09:34:52	S	Std
0.2 - 1	50.2	999	-	0.20007470A	PPB	26 Feb 2019 09:37:01	S	Std
2.5 - 1	52.5	11846	-	2.50007470A	PPB	26 Feb 2019 09:39:09	S	Std
5.0 - 1	55.0	23330	-	5.00007470A	PPB	26 Feb 2019 09:41:17	S	Std
7.5 - 1	57.5	34693	-	7.50007470A	PPB	26 Feb 2019 09:43:24	S	Std
10.0 - 1	510.0	46206	-	10.00007470A	PPB	26 Feb 2019 09:45:31	S	Std
ICV22 - 1	ICV22	18496	3.9792	-7470A	PPB	26 Feb 2019 09:57:23	U	SMPL
ICB22 - 1	ICB22	23	-0.0244	-7470A	PPB	26 Feb 2019 09:59:30	U	SMPL
CCV14 - 1	CCV14	23670	5.1006	-7470A	PPB	26 Feb 2019 10:01:36	U	SMPL
CCB14 - 1	CCB14	-55	-0.0413	-7470A	PPB	26 Feb 2019 10:03:42	U	SMPL
CRA - 1	CRA	988	0.1847	-7470A	PPB	26 Feb 2019 10:05:48	U	SMPL
HighStd - 1	HighStd	46776	10.1083	-7470A	PPB	26 Feb 2019 10:07:54	U	SMPL
ChkStd - 1	ChkStd	32133	6.9348	-7470A	PPB	26 Feb 2019 10:10:00	U	SMPL
PB117383BL - 1	PBW	-66	-0.0437	-7470A	PPB	26 Feb 2019 10:12:07	U	SMPL
PB117383BS - 1	LCSW	18424	3.9636	-7470A	PPB	26 Feb 2019 10:20:41	U	SMPL
K1348-04 - 1	OK-03-022219-A	9	-0.0274	-7470A	PPB	26 Feb 2019 10:22:47	U	SMPL
K1348-04DUP - 1	OK-03-022219-ADUP	41	-0.0205	-7470A	PPB	26 Feb 2019 10:24:55	U	SMPL
K1348-04MS - 1	OK-03-022219-AMS	16300	3.5033	-7470A	PPB	26 Feb 2019 10:27:03	U	SMPL
K1348-04MSD - 1	OK-03-022219-AMSD	16602	3.5687	-7470A	PPB	26 Feb 2019 10:29:11	U	SMPL
K1348-06 - 1	OK-03-022219-B	-27	-0.0352	-7470A	PPB	26 Feb 2019 10:31:17	U	SMPL
CCV15 - 1	CCV15	23392	5.0403	-7470A	PPB	26 Feb 2019 10:33:23	U	SMPL
CCB15 - 1	CCB15	20	-0.0251	-7470A	PPB	26 Feb 2019 10:35:29	U	SMPL
K1356-10 - 1	SU-02-022219-A	21	-0.0248	-7470A	PPB	26 Feb 2019 10:39:41	U	SMPL
K1356-12 - 1	SU-02-022219-B	9	-0.0274	-7470A	PPB	26 Feb 2019 10:41:47	U	SMPL
K1586-04 - 1	M-1	24	-0.0242	-7470A	PPB	26 Feb 2019 10:43:53	U	SMPL
K1592-04 - 1	TP-4	32	-0.0225	-7470A	PPB	26 Feb 2019 10:46:01	U	SMPL
K1595-02 - 1	TP-CBF-B10	40	-0.0207	-7470A	PPB	26 Feb 2019 10:48:08	U	SMPL
K1595-04 - 1	TP-CBF-B11	44	-0.0199	-7470A	PPB	26 Feb 2019 10:50:15	U	SMPL
K1595-06 - 1	TP-E-WALL-GRADE-BEAM	20	-0.0251	-7470A	PPB	26 Feb 2019 10:52:23	U	SMPL
K1595-08 - 1	TP-MH-C02	35	-0.0218	-7470A	PPB	26 Feb 2019 10:54:31	U	SMPL
K1597-02 - 1	TP-25	17	-0.0257	-7470A	PPB	26 Feb 2019 10:56:37	U	SMPL
CCV16 - 1	CCV16	23679	5.1025	-7470A	PPB	26 Feb 2019 10:58:44	U	SMPL
CCB16 - 1	CCB16	28	-0.0233	-7470A	PPB	26 Feb 2019 11:00:50	U	SMPL
K1597-04 - 1	TP-26	7	-0.0279	-7470A	PPB	26 Feb 2019 11:02:57	U	SMPL
K1597-12 - 1	TP-22	15	-0.0261	-7470A	PPB	26 Feb 2019 11:05:03	U	SMPL
K1597-16 - 1	TP-23	26	-0.0238	-7470A	PPB	26 Feb 2019 11:07:10	U	SMPL
PB117384BL - 1	PBW	60	-0.0164	-7470A	PPB	26 Feb 2019 11:09:16	U	SMPL
PB117384BS - 1	LCSW	16174	3.4760	-7470A	PPB	26 Feb 2019 11:11:23	U	SMPL
K1566-01 - 1	23377	-13	-0.0322	-7470A	PPB	26 Feb 2019 11:13:30	U	SMPL
K1567-01 - 1	COMP-1	31	-0.0227	-7470A	PPB	26 Feb 2019 11:15:38	U	SMPL
K1572-02 - 1	30019	31	-0.0227	-7470A	PPB	26 Feb 2019 11:17:46	U	SMPL
K1596-04 - 1	COMP-13	39	-0.0209	-7470A	PPB	26 Feb 2019 11:19:55	U	SMPL
K1596-05 - 1	COMP-14	62	-0.0160	-7470A	PPB	26 Feb 2019 11:22:02	U	SMPL
CCV17 - 1	CCV17	23653	5.0969	-7470A	PPB	26 Feb 2019 11:24:08	U	SMPL
CCB17 - 1	CCB17	-43	-0.0387	-7470A	PPB	26 Feb 2019 11:26:15	U	SMPL
K1596-06 - 1	COMP-15	20	-0.0251	-7470A	PPB	26 Feb 2019 11:28:21	U	SMPL
K1596-06DUP - 1	COMP-15DUP	56	-0.0173	-7470A	PPB	26 Feb 2019 11:30:27	U	SMPL
K1596-06MS - 1	COMP-15MS	17880	3.8457	-7470A	PPB	26 Feb 2019 11:36:58	U	SMPL
K1596-06MSD - 1	COMP-15MSD	17339	3.7285	-7470A	PPB	26 Feb 2019 11:39:04	U	SMPL
PB117399BL - 1	PBW	-9	-0.0313	-7470A	PPB	26 Feb 2019 11:41:11	U	SMPL
PB117399BS - 1	LCSW	16794	3.6104	-7470A	PPB	26 Feb 2019 11:43:19	U	SMPL
K1540-05 - 1	M1-T0-OZ(PRE)	113	-0.0049	-7470A	PPB	26 Feb 2019 11:45:26	U	SMPL
K1603-08 - 1	M1-T0-UV(PRE)	34	-0.0220	-7470A	PPB	26 Feb 2019 11:47:34	U	SMPL
K1603-15 - 1	M2-T0-UV(PRE)	52	-0.0181	-7470A	PPB	26 Feb 2019 11:49:42	U	SMPL
K1603-15DUP - 1	M2-T0-UV(PRE)DUP	53	-0.0179	-7470A	PPB	26 Feb 2019 11:51:48	U	SMPL
CCV18 - 1	CCV18	23864	5.1426	-7470A	PPB	26 Feb 2019 11:53:55	U	SMPL
CCB18 - 1	CCB18	-23	-0.0344	-7470A	PPB	26 Feb 2019 11:56:02	U	SMPL
K1603-15MS - 1	M2-T0-UV(PRE)MS	16818	3.6156	-7470A	PPB	26 Feb 2019 11:58:08	U	SMPL
K1603-15MSD - 1	M2-T0-UV(PRE)MSD	16775	3.6062	-7470A	PPB	26 Feb 2019 12:00:15	U	SMPL
K1605-01 - 1	N48966	5669	1.1992	-7470A	PPB	26 Feb 2019 12:02:21	U	SMPL
PB117383TB - 1	PB117310TB	0	-0.0294	-7470A	PPB	26 Feb 2019 12:04:28	U	SMPL
PB117384TB - 1	PB117333TB	26	-0.0238	-7470A	PPB	26 Feb 2019 12:06:34	U	SMPL
K1348-04LX5 - 1		3	-0.0287	-7470A	PPB	26 Feb 2019 12:08:42	U	SMPL
K1348-04A - 1		17127	3.6825	-7470A	PPB	26 Feb 2019 12:10:49	U	SMPL
K1596-06LX5 - 1		-28	-0.0355	-7470A	PPB	26 Feb 2019 12:12:57	U	SMPL
K1596-06A - 1		17749	3.8173	-7470A	PPB	26 Feb 2019 12:15:05	U	SMPL
K1603-15LX5 - 1		-47	-0.0396	-7470A	PPB	26 Feb 2019 12:17:11	U	SMPL
CCV19 - 1	CCV19	24092	5.1920	-7470A	PPB	26 Feb 2019 12:19:18	U	SMPL
CCB19 - 1	CCB19	-2	-0.0298	-7470A	PPB	26 Feb 2019 12:21:26	U	SMPL
K1603-15A - 1		16918	3.6372	-7470A	PPB	26 Feb 2019 12:23:33	U	SMPL
CCV20 - 1	CCV20	23868	5.1435	-7470A	PPB	26 Feb 2019 12:25:40	U	SMPL
CCB20 - 1	CCB20	-30	-0.0359	-7470A	PPB	26 Feb 2019 12:27:47	U	SMPL