



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100111

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	01/02/2019	12:44
2	J6589-01		1	8	>100.0	>212.0	>212.0	01/02/2019	12:51
3	J6589-01DUP		1	8	>100.0	>212.0	>212.0	01/02/2019	13:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB100112

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp010219

WorkList ID : 120914

Date : 1/2/2019 12:22:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2019	Solid	J6589-01	Flash Point	Cool 4 deg C	PSEG01	E31	13178	12/28/2018	1010A

Date/Time 01/02/2019 12:30
Received by: AK
Relinquished by: CP

Date/Time 01/02/18 13:20
Received by: CP
Relinquished by: AK