



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100112

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	01/02/2019	13:15
2	J6587-01		1	6	>100.0	>212.0	>212.0	01/02/2019	13:22
3	J6587-01DUP		1	6	>100.0	>212.0	>212.0	01/02/2019	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

03/02/19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp010219-2

WorkList ID : 120915

Date : 1/2/2019 12:22:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/03/2019	Water	J6587-01	Flash Point	Cool 4 deg C	PARS02		WC-20181227	12/27/2018	1010A

Date/Time 01/02/2019 13:10
 Received by: AV
 Relinquished by: Q

Date/Time 01/02/2019 13:40
 Received by: Cy
 Relinquished by: de