



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100113

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	01/02/2019	12:40
2	CAL2	1	Water			20.1	7.00	01/02/2019	12:41
3	CAL3	1	Water			20.1	10.02	01/02/2019	12:42
4	ICV	1	Water			20.1	6.99	01/02/2019	12:43
5	CCV1	1	Water			20.1	2.01	01/02/2019	12:45
6	K1020-01	1	Water			19.2	4.52	01/02/2019	12:47
7	K1020-01DUP	1	Water			19.3	4.53	01/02/2019	12:48
8	K1020-02	1	Water			19.3	4.39	01/02/2019	12:49
9	K1020-03	1	Water			19.1	8.98	01/02/2019	12:50
10	CCV2	1	Water			20.2	12.02	01/02/2019	13:00

2-B 100113

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1020

WorkList ID : 120916

Date : 1/2/2019 12:34:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/27/2018	Water	K1020-01 - ☒	pH	Cool 4 deg C	STER05	D21	BOTTLE-1	12/27/2018	SM 4500-H B
12/27/2018	Water	K1020-02 - ☒	pH	Cool 4 deg C	STER05	D21	BOTTLE-2	12/27/2018	SM 4500-H B
12/28/2018	Water	K1020-03 - ☒	pH	Cool 4 deg C	STER05	D21	BOTTLE-3	12/28/2018	SM 4500-H B

Date/Time 01-02-19 12:35 PM

Received by: JoRelinquished by: CS

Date/Time 01-02-19 1:50 PM

Received by: CSRelinquished by: Jo