



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100116

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	01/02/2019	17:10
2	CAL2	1	Water			20.2	7.00	01/02/2019	17:11
3	CAL3	1	Water			20.2	10.03	01/02/2019	17:12
4	ICV	1	Water			20.2	7.01	01/02/2019	17:14
5	CCV1	1	Water			20.1	2.02	01/02/2019	17:15
6	K1002-02	1	Solid	20.02	20	20.5	6.96	01/02/2019	17:16
7	K1002-02DUP	1	Solid	20.03	20	20.5	6.97	01/02/2019	17:18
8	K1002-04	1	Solid	20.04	20	20.3	7.53	01/02/2019	17:20
9	K1002-06	1	Solid	20.01	20	20.4	6.48	01/02/2019	17:21
10	K1002-08	1	Solid	20.04	20	20.4	7.88	01/02/2019	17:22
11	K1016-02	1	Solid	20.03	20	20.5	7.95	01/02/2019	17:24
12	K1017-04	1	Solid	20.03	20	20.1	8.23	01/02/2019	17:25
13	K1017-08	1	Solid	20.05	20	20.3	7.25	01/02/2019	17:26
14	K1017-12	1	Solid	20.03	20	20.3	7.46	01/02/2019	17:29
15	K1017-16	1	Solid	20.04	20	20.5	7.69	01/02/2019	17:30
16	CCV2	1	Water			20.2	12.02	01/02/2019	17:31
17	K1017-20	1	Solid	20.02	20	20.4	7.50	01/02/2019	17:33
18	K1017-24	1	Solid	20.04	20	20.2	7.18	01/02/2019	17:34
19	K1019-02	1	Solid	20.03	20	20.1	6.77	01/02/2019	17:35
20	K1019-04	1	Solid	20.02	20	20.2	7.30	01/02/2019	17:36

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	K1019-06	1	Solid	20.03	20	20.4	8.93	01/02/2019	17:38
22	K1019-08	1	Solid	20.02	20	20.3	7.27	01/02/2019	17:39
23	CCV3	1	Water			20.1	2.01	01/02/2019	17:40

7B 100118

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivityW K1023

WorkList ID : 120926

Date : 1/2/2019 4:31:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2019	Water	K1017-26	Corrosivity	Cool 4 deg C	PSEG01	E61	TP-5	01/02/2019	9040C
01/02/2019	Water	K1023-01	Corrosivity	Cool 4 deg C	SEAC01	D62	PURGE-WATER	01/02/2019	9040C

Date/Time 01-02-19 2:10 PM
Received by: SO
Relinquished by: SO

Date/Time 01-02-19 6:30 PM
Received by: SO
Relinquished by: SO