



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100118

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	01/02/2019	17:10
2	CAL2	1	Water			20.1	7.00	01/02/2019	17:11
3	CAL3	1	Water			20.2	10.03	01/02/2019	17:12
4	ICV	1	Water			20.2	7.01	01/02/2019	17:14
5	CCV1	1	Water			20.1	2.01	01/02/2019	17:40
6	K1017-25	1	Water			20.1	6.70	01/02/2019	17:41
7	K1017-25DUP	1	Water			20.2	6.71	01/02/2019	17:42
8	CCV2	1	Water			20.2	12.02	01/02/2019	17:45

2B 100118

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1023

WorkList ID : 120927

Date : 1/2/2019 4:32:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2019	Water	K1017-25	pH	Cool 4 deg C	PSEG01	E61	TP-5	01/02/2019	9040C

Date/Time 01-02-19 3:10 PM
Received by: SN
Relinquished by: SN

Date/Time 01-02-19 6:32 PM
Received by: SN
Relinquished by: SN