

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 01/02/2019 15:30 OUT: 01/03/2019 07:00

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 01/03/2019 08:05 OUT: 01/03/2019 09:40

Run Number: LB100124

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB100124BL	LB100124BL	82.4569	82.4569	100	82.4569	82.4569	82.4569	0.0000	0
2	K1018-01	TEST-PIT-13	90.0657	90.0657	100	90.0767	90.0767	90.0767	0.0110	110
3	K1018-02	TEST-PIT-D11-12	94.2679	94.2679	100	94.3662	94.3662	94.3662	0.0983	983
4	K1018-02DUP	TEST-PIT-D11-12DUP	83.4405	83.4405	100	83.5380	83.5380	83.5380	0.0975	975

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS010219

WorkList ID : 120922

Date : 1/2/2019 2:31:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/09/2019	Water	K1018-01 H	TS	Cool 4 deg C	TULL02	D61	TEST-PIT-13	01/02/2019	SM2540 B
01/09/2019	Water	K1018-02 H	TS	Cool 4 deg C	TULL02	D61	TEST-PIT-D11&12	01/02/2019	SM2540 B

Date/Time 01/02/18 14:20
 Received by: AS
 Relinquished by: [Signature]

Date/Time 01/02/18 15:45
 Received by: [Signature]
 Relinquished by: [Signature]