

LB 100139


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

48597



A= 0.0000e+000

B= 2.0555e-004

C= -2.6632e-002

Rho= 0.9999132

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	64	0.000	64				
0.2	0.200	0.184	-0.016	1026	0.0 %	1026				
2.5	2.500	2.559	0.059	12577	0.0 %	12577				
5.0	5.000	4.937	-0.063	24150	0.0 %	24150				
7.5	7.500	7.571	0.071	36961	0.0 %	36961				
10.0	10.000	9.963	-0.037	48597	0.0 %	48597				

