


7470A

Linear

 μ Abs.:

38503



A= 0.0000e+000

B= 2.5845e-004

C= 5.8378e-003

Rho= 0.9998978

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.060	-0.060	-256	0.000	-256				
0.2	0.200	0.172	-0.028	643	0.0 %	643				
2.5	2.500	2.597	0.097	10024	0.0 %	10024				
5.0	5.000	5.033	0.033	19452	0.0 %	19452				
7.5	7.500	7.501	0.001	29002	0.0 %	29002				
10.0	10.000	9.957	-0.043	38503	0.0 %	38503	0			

LB100151 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jan 2019 09:15:35

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-256	-	0.00007470A	PPB		04 Jan 2019 09:16:47	S	Std
0.2 - 1	50.2	643	-	0.20007470A	PPB		04 Jan 2019 09:18:57	S	Std
2.5 - 1	52.5	10024	-	2.50007470A	PPB		04 Jan 2019 09:21:08	S	Std
5.0 - 1	55.0	19452	-	5.00007470A	PPB		04 Jan 2019 09:23:17	S	Std
7.5 - 1	57.5	29002	-	7.50007470A	PPB		04 Jan 2019 09:25:27	S	Std
10.0 - 1	510.0	38503	-	10.00007470A	PPB		04 Jan 2019 09:27:34	S	Std
ICV51 - 1	ICV51	15328	3.9674	-7470A	PPB		04 Jan 2019 09:30:29	U	SMPL
ICB51 - 1	ICB51	-11	0.0030	-7470A	PPB		04 Jan 2019 09:32:35	U	SMPL
CCV58 - 1	CCV58	18680	4.8337	-7470A	PPB		04 Jan 2019 09:34:42	U	SMPL
CCB58 - 1	CCB58	-16	0.0017	-7470A	PPB		04 Jan 2019 09:36:48	U	SMPL
CRA - 1	CRA	820	0.2178	-7470A	PPB		04 Jan 2019 09:38:55	U	SMPL
HighStd - 1	HighStd	38412	9.9335	-7470A	PPB		04 Jan 2019 09:41:02	U	SMPL
ChkStd - 1	ChkStd	27254	7.0497	-7470A	PPB		04 Jan 2019 09:43:09	U	SMPL
PB116146BL - 1	PBW	23	0.0118	-7470A	PPB		04 Jan 2019 09:45:18	U	SMPL
PB116146BS - 1	LCSW	13591	3.5185	-7470A	PPB		04 Jan 2019 09:47:25	U	SMPL
PB116146TB - 1	PB116104TB	97	0.0309	-7470A	PPB		04 Jan 2019 09:49:34	U	SMPL
K1002-02 - 1	AU-05-010219-A	4	0.0069	-7470A	PPB		04 Jan 2019 09:53:51	U	SMPL
K1002-02DUP - 1	AU-05-010219-ADUP	10	0.0084	-7470A	PPB		04 Jan 2019 10:07:56	U	SMPL
K1002-02MS - 1	AU-05-010219-AMS	14443	3.7387	-7470A	PPB		04 Jan 2019 10:10:31	U	SMPL
K1002-02MSD - 1	AU-05-010219-AMSD	14548	3.7658	-7470A	PPB		04 Jan 2019 10:12:37	U	SMPL
CCV59 - 1	CCV59	19261	4.9839	-7470A	PPB		04 Jan 2019 10:14:44	U	SMPL
CCB59 - 1	CCB59	-25	-0.0006	-7470A	PPB		04 Jan 2019 10:16:51	U	SMPL
K1002-04 - 1	AU-05-010219-B	2	0.0064	-7470A	PPB		04 Jan 2019 10:18:57	U	SMPL
K1002-06 - 1	AU-05-010219-C	149	0.0443	-7470A	PPB		04 Jan 2019 10:21:04	U	SMPL
K1002-08 - 1	AU-05-010219-D	354	0.0973	-7470A	PPB		04 Jan 2019 10:23:12	U	SMPL
K1004-01 - 1	BS01-TO-BS09-B2	557	0.1498	-7470A	PPB		04 Jan 2019 10:25:19	U	SMPL
K1016-02 - 1	TR-04-010219	87	0.0283	-7470A	PPB		04 Jan 2019 10:27:27	U	SMPL
K1017-04 - 1	TP-1	85	0.0278	-7470A	PPB		04 Jan 2019 10:29:35	U	SMPL
K1017-08 - 1	TP-2	24	0.0120	-7470A	PPB		04 Jan 2019 10:31:44	U	SMPL
K1017-12 - 1	TP-3	14	0.0095	-7470A	PPB		04 Jan 2019 10:33:52	U	SMPL
K1017-16 - 1	TP-4	14	0.0095	-7470A	PPB		04 Jan 2019 10:36:00	U	SMPL
K1017-20 - 1	TP-5	27	0.0128	-7470A	PPB		04 Jan 2019 10:38:07	U	SMPL
CCV60 - 1	CCV60	18952	4.9040	-7470A	PPB		04 Jan 2019 10:40:15	U	SMPL
CCB60 - 1	CCB60	42	0.0167	-7470A	PPB		04 Jan 2019 10:42:21	U	SMPL
K1017-24 - 1	TP-6	25	0.0123	-7470A	PPB		04 Jan 2019 10:44:29	U	SMPL
K1019-02 - 1	TEST-PIT-D-11-12-A	17	0.0102	-7470A	PPB		04 Jan 2019 10:46:36	U	SMPL
K1019-04 - 1	TEST-PIT-D-11-12-B	11	0.0087	-7470A	PPB		04 Jan 2019 10:48:43	U	SMPL
K1019-06 - 1	A-GLICAL-PILE	88	0.0286	-7470A	PPB		04 Jan 2019 10:50:51	U	SMPL
K1019-08 - 1	TEST-PIT-D13	32	0.0141	-7470A	PPB		04 Jan 2019 10:52:58	U	SMPL
PB116147BL - 1	PBW	119	0.0366	-7470A	PPB		04 Jan 2019 10:55:07	U	SMPL
PB116147BS - 1	LCSW	13635	3.5298	-7470A	PPB		04 Jan 2019 10:57:15	U	SMPL
K1017-25 - 1	TP-5	2395	0.6248	-7470A	PPB		04 Jan 2019 10:59:24	U	SMPL
K1018-01 - 1	TEST-PIT-13	-87	-0.0166	-7470A	PPB		04 Jan 2019 11:01:31	U	SMPL
K1018-02 - 1	TEST-PIT-D11-12	106	0.0332	-7470A	PPB		04 Jan 2019 11:03:39	U	SMPL
CCV61 - 1	CCV61	19495	5.0444	-7470A	PPB		04 Jan 2019 11:05:46	U	SMPL
CCB61 - 1	CCB61	44	0.0172	-7470A	PPB		04 Jan 2019 11:07:54	U	SMPL
K1023-01 - 1	PURGE-WATER	-28	-0.0014	-7470A	PPB		04 Jan 2019 11:10:01	U	SMPL
K1029-01 - 1	B12_123118	5	0.0071	-7470A	PPB		04 Jan 2019 11:12:09	U	SMPL
K1029-01DUP - 1	B12_123118DUP	-19	0.0009	-7470A	PPB		04 Jan 2019 11:14:16	U	SMPL
K1029-01MS - 1	B12_123118MS	15867	4.1067	-7470A	PPB		04 Jan 2019 11:16:24	U	SMPL
K1029-01MSD - 1	B12_123118MSD	16029	4.1486	-7470A	PPB		04 Jan 2019 11:18:32	U	SMPL
K1029-02 - 1	B12_123118	-87	-0.0166	-7470A	PPB		04 Jan 2019 11:20:41	U	SMPL
K1029-03 - 1	B1_123118	-37	-0.0037	-7470A	PPB		04 Jan 2019 11:22:49	U	SMPL
K1029-04 - 1	B1_123118	-34	-0.0029	-7470A	PPB		04 Jan 2019 11:24:58	U	SMPL
K1029-05 - 1	MW12_123118	139	0.0418	-7470A	PPB		04 Jan 2019 11:27:06	U	SMPL
K1029-06 - 1	MW12_123118	119	0.0366	-7470A	PPB		04 Jan 2019 11:29:13	U	SMPL
CCV62 - 1	CCV62	19364	5.0105	-7470A	PPB		04 Jan 2019 11:31:21	U	SMPL
CCB62 - 1	CCB62	72	0.0244	-7470A	PPB		04 Jan 2019 11:33:28	U	SMPL
J6587-06 - 1	WC-20181227	221	0.0630	-7470A	PPB		04 Jan 2019 11:35:36	U	SMPL
K1002-02LX5 - 1		44	0.0172	-7470A	PPB		04 Jan 2019 11:37:43	U	SMPL
K1002-02A - 1		15733	4.0721	-7470A	PPB		04 Jan 2019 11:39:51	U	SMPL
K1029-01LX5 - 1		-3	0.0051	-7470A	PPB		04 Jan 2019 11:41:58	U	SMPL
K1029-01A - 1		15930	4.1230	-7470A	PPB		04 Jan 2019 11:44:06	U	SMPL
CCV63 - 1	CCV63	19482	5.0410	-7470A	PPB		04 Jan 2019 11:46:14	U	SMPL
CCB63 - 1	CCB63	66	0.0229	-7470A	PPB		04 Jan 2019 11:48:23	U	SMPL