



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100158

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/03/2019	16:15
2	CAL2	1	Water	NA	NA	20.1	6.99	01/03/2019	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/03/2019	16:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/03/2019	16:18
5	CCV1	1	Water	NA	NA	20.1	2.01	01/03/2019	16:20
6	K1040-03	1	Solid	20.02	20	21.0	7.76	01/03/2019	16:21
7	K1041-04	1	Solid	20.03	20	21.4	8.64	01/03/2019	16:23
8	K1041-04DUP	1	Solid	20.04	20	21.3	8.65	01/03/2019	16:25
9	CCV2	1	Water	NA	NA	20.1	12.01	01/03/2019	16:26

VB 100158

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity K1040

WorkList ID : 121020

Date : 1/3/2019 4:02:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/03/2019	Solid	K1040-03	Corrosivity	Cool 4 deg C	PSEG01	E61	RBR-200030-RBR-200028-RT	01/03/2019	9045D
01/03/2019	Solid	K1041-04	Corrosivity	Cool 4 deg C	PSEG01	E61	OS-1-5-COMP	01/03/2019	9045D

Date/Time 01-03-19 4:10 PM
 Received by: SO
 Relinquished by: SO

Date/Time 01-03-19 4:00 PM
 Received by: SO
 Relinquished by: SO