



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100160

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/03/2019	16:15
2	CAL2	1	Water	NA	NA	20.1	6.99	01/03/2019	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/03/2019	16:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/03/2019	16:18
5	CCV1	1	Water	NA	NA	20.1	2.01	01/03/2019	16:30
6	K1042-01	1	Water	NA	NA	16.2	6.53	01/03/2019	16:31
7	K1042-01DUP	1	Water	NA	NA	16.3	6.54	01/03/2019	16:32
8	K1043-01	1	Water	NA	NA	16.5	6.95	01/03/2019	16:33
9	CCV2	1	Water	NA	NA	20.1	12.02	01/03/2019	16:34

WORKLIST(Hardcopy Internal Chain)

LB 100160

WorkList Name : PH K1042

WorkList ID : 121022

Date : 1/3/2019 17:10:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/03/2019	Water	K1042-01	pH	Cool 4 deg C	PSEG01	E61	JC-1319-WS	01/03/2019	9040C
01/03/2019	Water	K1043-01	pH	Cool 4 deg C	PSEG01	E61	HB-1319-WS	01/03/2019	9040C

Date/Time 01-03-19 17:15 PM
 Received by: SP
 Relinquished by: CO

Date/Time 01-03-19 6:00 PM
 Received by: CO
 Relinquished by: SP