

LB100161


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54062



A= 0.0000e+000

B= 1.8573e-004

C= -3.8462e-002

Rho= 0.9999967

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	239	0.000	239				
0.2	0.200	0.195	-0.005	1255	0.0 %	1255				
2.5	2.500	2.492	-0.008	13625	0.0 %	13625				
5.0	5.000	5.016	0.016	27216	0.0 %	27216				
7.5	7.500	7.488	-0.012	40525	0.0 %	40525				
10.0	10.000	10.003	0.003	54062	0.0 %	54062				

LB100161 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jan 2019 12:18:28

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50-0	239	-	0.0000	7471B	PPB	04 Jan 2019 12:21:20	S	Std
0.2 - 1	50-2	1255	-	0.2000	7471B	PPB	04 Jan 2019 12:23:38	S	Std
2.5 - 1	52.5	13625	-	2.5000	7471B	PPB	04 Jan 2019 12:25:56	S	Std
5.0 - 1	55.0	27216	-	5.0000	7471B	PPB	04 Jan 2019 12:28:14	S	Std
7.5 - 1	57.5	40525	-	7.5000	7471B	PPB	04 Jan 2019 12:30:32	S	Std
10.0 - 1	510-0	54062	-	10.0000	7471B	PPB	04 Jan 2019 12:32:49	S	Std
ICV52 - 1	ICV52	22128	4.0714	-	7471B	PPB	04 Jan 2019 12:38:01	U	SMPL
ICB52 - 1	ICB52	143	-0.0119	-	7471B	PPB	04 Jan 2019 12:40:16	U	SMPL
CCV64 - 1	CCV64	28150	5.1899	-	7471B	PPB	04 Jan 2019 12:42:32	U	SMPL
CCB64 - 1	CCB64	131	-0.0141	-	7471B	PPB	04 Jan 2019 12:44:49	U	SMPL
CRA - 1	CRA	1262	0.1959	-	7471B	PPB	04 Jan 2019 12:47:05	U	SMPL
HighStd - 1	HighStd	53898	9.9721	-	7471B	PPB	04 Jan 2019 12:49:21	U	SMPL
ChkStd - 1	ChkStd	37329	6.8947	-	7471B	PPB	04 Jan 2019 12:51:37	U	SMPL
PB116161BL - 1	PBS	-90	-0.0552	-	7471B	PPB	04 Jan 2019 12:53:55	U	SMPL
PB116161BS - 1	LCSS	19215	3.5304	-	7471B	PPB	04 Jan 2019 12:56:14	U	SMPL
J6562-01 - 1	P001-SS072-0002-01	4318	0.7635	-	7471B	PPB	04 Jan 2019 12:58:31	U	SMPL
J6562-01DUP - 1	P001-SS072-0002-01DUP	4324	0.7646	-	7471B	PPB	04 Jan 2019 13:00:49	U	SMPL
J6562-01MS - 1	P001-SS072-0002-01MS	27224	5.0179	-	7471B	PPB	04 Jan 2019 13:03:07	U	SMPL
J6562-01MSD - 1	P001-SS072-0002-01MSD	26284	4.8433	-	7471B	PPB	04 Jan 2019 13:05:25	U	SMPL
J6562-02 - 1	P001-SS075-0206-01	10671	1.9435	-	7471B	PPB	04 Jan 2019 13:07:42	U	SMPL
CCV65 - 1	CCV65	28018	5.1654	-	7471B	PPB	04 Jan 2019 13:09:59	U	SMPL
CCB65 - 1	CCB65	118	-0.0165	-	7471B	PPB	04 Jan 2019 13:12:16	U	SMPL
J6562-03 - 1	P001-SS077-0206-01	5456	0.9749	-	7471B	PPB	04 Jan 2019 13:14:33	U	SMPL
J6562-04 - 1	P001-SS079-0002-01	13099	2.3944	-	7471B	PPB	04 Jan 2019 13:16:49	U	SMPL
J6562-05 - 1	P001-SS079-0206-01	9304	1.6896	-	7471B	PPB	04 Jan 2019 13:19:05	U	SMPL
J6562-06 - 1	P001-SS080-0002-01	61496	11.3833	-	7471B	PPB	04 Jan 2019 13:21:22	U	SMPL
J6562-07 - 1	P001-SS080-0612-01	12829	2.3443	-	7471B	PPB	04 Jan 2019 13:23:39	U	SMPL
J6562-08 - 1	P001-SS081-0002-01	56666	10.4862	-	7471B	PPB	04 Jan 2019 13:25:58	U	SMPL
J6562-09 - 1	P001-SS082-1218-01	2989	0.5167	-	7471B	PPB	04 Jan 2019 13:28:15	U	SMPL
J6562-10 - 1	P001-SS083-0612-01	12531	2.2889	-	7471B	PPB	04 Jan 2019 13:30:35	U	SMPL
J6562-12 - 1	P001-SS085-0206-01	11059	2.0155	-	7471B	PPB	04 Jan 2019 13:32:53	U	SMPL
J6562-13 - 1	P001-SS085-0612-01	5337	0.9528	-	7471B	PPB	04 Jan 2019 13:35:09	U	SMPL
CCV66 - 1	CCV66	28079	5.1767	-	7471B	PPB	04 Jan 2019 13:37:26	U	SMPL
CCB66 - 1	CCB66	119	-0.0164	-	7471B	PPB	04 Jan 2019 13:39:43	U	SMPL
J6562-14 - 1	P001-SS086-0002-01	3743	0.6567	-	7471B	PPB	04 Jan 2019 13:42:00	U	SMPL
J6562-15 - 1	P001-SS087-0002-01	4769	0.8473	-	7471B	PPB	04 Jan 2019 13:44:16	U	SMPL
J6562-16 - 1	P001-SS087-0206-01	11614	2.1186	-	7471B	PPB	04 Jan 2019 13:46:33	U	SMPL
J6562-17 - 1	P001-SS088-0612-01	3596	0.6294	-	7471B	PPB	04 Jan 2019 13:48:49	U	SMPL
J6562-18 - 1	P001-SS089-0206-01	8358	1.5139	-	7471B	PPB	04 Jan 2019 13:51:06	U	SMPL
J6562-19 - 1	P001-SS090-0612-01	9875	1.7956	-	7471B	PPB	04 Jan 2019 13:53:23	U	SMPL
J6562-20 - 1	P001-SS176-1218-01	8133	1.4721	-	7471B	PPB	04 Jan 2019 13:55:41	U	SMPL
PB116162BL - 1	PBS	-1	-0.0386	-	7471B	PPB	04 Jan 2019 13:57:58	U	SMPL
PB116162BS - 1	LCSS	19508	3.5848	-	7471B	PPB	04 Jan 2019 14:00:16	U	SMPL
J6563-04 - 1	P001-SS064-0002-01	9298	1.6885	-	7471B	PPB	04 Jan 2019 14:02:33	U	SMPL
CCV67 - 1	CCV67	27869	5.1377	-	7471B	PPB	04 Jan 2019 14:04:50	U	SMPL
CCB67 - 1	CCB67	129	-0.0145	-	7471B	PPB	04 Jan 2019 14:07:06	U	SMPL
J6563-04DUP - 1	P001-SS064-0002-01DUP	8446	1.5302	-	7471B	PPB	04 Jan 2019 14:12:08	U	SMPL
J6563-05 - 1	J6563-04MS	29424	5.4265	-	7471B	PPB	04 Jan 2019 14:14:24	U	SMPL
J6563-06 - 1	J6563-04MSD	29424	5.4265	-	7471B	PPB	04 Jan 2019 14:16:41	U	SMPL
J6563-07 - 1	P001-SS064-0002-02	8567	1.5527	-	7471B	PPB	04 Jan 2019 14:18:58	U	SMPL
J6563-08 - 1	P001-SS065-0002-01	12329	2.2514	-	7471B	PPB	04 Jan 2019 14:21:16	U	SMPL
J6563-09 - 1	P001-SS066-0206-01	6004	1.0767	-	7471B	PPB	04 Jan 2019 14:23:33	U	SMPL
J6563-10 - 1	P001-SS067-0002-01	10220	1.8597	-	7471B	PPB	04 Jan 2019 14:25:52	U	SMPL
J6563-11 - 1	P001-SS068-0206-01	9339	1.6961	-	7471B	PPB	04 Jan 2019 14:28:10	U	SMPL
J6563-12 - 1	P001-SS069-0002-01	6706	1.2071	-	7471B	PPB	04 Jan 2019 14:30:28	U	SMPL
J6563-13 - 1	P001-SS070-0206-01	10862	1.9790	-	7471B	PPB	04 Jan 2019 14:32:45	U	SMPL
CCV68 - 1	CCV68	28015	5.1648	-	7471B	PPB	04 Jan 2019 14:35:01	U	SMPL
CCB68 - 1	CCB68	115	-0.0171	-	7471B	PPB	04 Jan 2019 14:37:18	U	SMPL
J6563-14 - 1	J6563-13MS	32001	5.9051	-	7471B	PPB	04 Jan 2019 14:39:36	U	SMPL
J6563-15 - 1	J6563-13MSD	30044	5.5417	-	7471B	PPB	04 Jan 2019 14:41:52	U	SMPL
J6564-01 - 1	P001-SS070-0206-02	11690	2.1327	-	7471B	PPB	04 Jan 2019 14:44:09	U	SMPL
J6564-10 - 1	P001-SS109-0206-02	12049	2.1994	-	7471B	PPB	04 Jan 2019 14:46:27	U	SMPL
J6564-12 - 1	P001-SS115-0206-02	23919	4.4041	-	7471B	PPB	04 Jan 2019 14:48:44	U	SMPL
J6564-15 - 1	P001-SS140-0002-02	246588	45.7608	-	7471B	PPB	04 Jan 2019 14:51:01	U	SMPL
J6564-16 - 1	P001-SS173-0206-02	13879	2.5393	-	7471B	PPB	04 Jan 2019 14:53:20	U	SMPL
J6564-17 - 1	P001-SS183-1218-02	1440	0.2290	-	7471B	PPB	04 Jan 2019 14:55:49	U	SMPL
J6564-18 - 1	P001-SS184-0206-02	10604	1.9310	-	7471B	PPB	04 Jan 2019 14:58:08	U	SMPL
J6564-19 - 1	P001-SS193-0612-02	5863	1.0505	-	7471B	PPB	04 Jan 2019 15:00:25	U	SMPL
CCV69 - 1	CCV69	27882	5.1401	-	7471B	PPB	04 Jan 2019 15:02:42	U	SMPL
CCB69 - 1	CCB69	107	-0.0186	-	7471B	PPB	04 Jan 2019 15:04:59	U	SMPL
PB116163BL - 1	PBS	202	-0.0009	-	7471B	PPB	04 Jan 2019 15:12:29	U	SMPL
PB116163BS - 1	LCSS	20401	3.7507	-	7471B	PPB	04 Jan 2019 15:14:45	U	SMPL
K1002-01 - 1	AU-05-010219-A	2771	0.4762	-	7471B	PPB	04 Jan 2019 15:17:02	U	SMPL
K1002-01DUP - 1	AU-05-010219-ADUP	2913	0.5026	-	7471B	PPB	04 Jan 2019 15:19:19	U	SMPL
K1002-01MS - 1	AU-05-010219-AMS	22274	4.0985	-	7471B	PPB	04 Jan 2019 15:21:36	U	SMPL
K1002-01MSD - 1	AU-05-010219-AMSD	22001	4.0478	-	7471B	PPB	04 Jan 2019 15:23:53	U	SMPL
K1002-03 - 1	AU-05-010219-B	1699	0.2771	-	7471B	PPB	04 Jan 2019 15:26:11	U	SMPL
K1002-05 - 1	AU-05-010219-C	3442	0.6008	-	7471B	PPB	04 Jan 2019 15:28:29	U	SMPL
K1002-07 - 1	AU-05-010219-D	3086	0.5347	-	7471B	PPB	04 Jan 2019 15:30:48	U	SMPL
K1004-02 - 1	BS01-TO-BS09-B2	13081	2.3911	-	7471B	PPB	04 Jan 2019 15:33:05	U	SMPL
CCV70 - 1	CCV70	28324	5.2222	-	7471B	PPB	04 Jan 2019 15:35:22	U	SMPL
CCB70 - 1	CCB70	118	-0.0165	-	7471B	PPB	04 Jan 2019 15:37:41	U	SMPL
K1016-01 - 1	TR-04-010219	3434	0.5993	-	7471B	PPB	04 Jan 2019 15:39:59	U	SMPL
K1017-01 - 1	TP-1	20592	3.7861	-	7471B	PPB	04 Jan 2019 15:42:16	U	SMPL
K1017-05 - 1	TP-2	10791	1.9658	-	7471B	PPB	04 Jan 2019 15:44:33	U	SMPL
K1017-09 - 1	TP-3	2284	0.3857	-	7471B	PPB	04 Jan 2019 15:46:51	U	SMPL
K1017-13 - 1	TP-4	7451	1.3454	-	7471B	PPB	04 Jan 2019 15:49:08	U	SMPL
K1017-17 - 1	TP-5	5352	0.9556	-	7471B	PPB	04 Jan 2019 15:51:26	U	SMPL
K1017-21 - 1	TP-6	9349	1.6979	-	7471B	PPB	04 Jan 2019 15:53:44	U	SMPL
K1019-01 - 1	TEST-PIT-D-11-12-A	6173	1.1081	-	7471B	PPB	04 Jan 2019 15:56:01	U	SMPL
K1019-03 - 1	TEST-PIT-D-11-12-B	2507	0.4272	-	7471B	PPB	04 Jan 2019 15:58:19	U	SMPL
K1019-05 - 1	A-GLICAL-PILE	22300	4.1034	-	7471B	PPB	04 Jan 2019 16:00:36	U	SMPL
CCV71 - 1	CCV71	28207	5.2005	-	7471B	PPB	04 Jan 2019 16:02:53	U	SMPL

LB100161 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jan 2019 12:18:28

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
CCB71 - 1	CCB71	105	-0.0190	-7471B	PPB		04 Jan 2019 16:05:11	U	SMPL
K1019-07 - 1	TEST-PIT-D13	7090	1.2784	-7471B	PPB		04 Jan 2019 16:07:29	U	SMPL
K1040-01 - 1	RBR-200030-RBR-200028-RT-3483	616	0.0759	-7471B	PPB		04 Jan 2019 16:09:47	U	SMPL
K1041-01 - 1	OS-1-5-COMP	143	-0.0119	-7471B	PPB		04 Jan 2019 16:12:04	U	SMPL
J6562-01LX5 - 1		975	0.1426	-7471B	PPB		04 Jan 2019 16:14:21	U	SMPL
J6562-01A - 1		27611	5.0898	-7471B	PPB		04 Jan 2019 16:16:38	U	SMPL
J6563-04LX5 - 1		1687	0.2749	-7471B	PPB		04 Jan 2019 16:18:55	U	SMPL
J6563-04A - 1		31779	5.8639	-7471B	PPB		04 Jan 2019 16:21:14	U	SMPL
K1002-01LX5 - 1		625	0.0776	-7471B	PPB		04 Jan 2019 16:23:31	U	SMPL
K1002-01A - 1		25683	4.7317	-7471B	PPB		04 Jan 2019 16:25:50	U	SMPL
CCV72 - 1	CCV72	28202	5.1995	-7471B	PPB		04 Jan 2019 16:28:07	U	SMPL
CCB72 - 1	CCB72	99	-0.0201	-7471B	PPB		04 Jan 2019 16:30:26	U	SMPL
J6562-06DLX2 - 1		31457	5.8041	-7471B	PPB		04 Jan 2019 16:32:43	U	SMPL
J6562-08DLX2 - 1		28962	5.3407	-7471B	PPB		04 Jan 2019 16:35:00	U	SMPL
J6564-15DLX10 - 1		26936	4.9644	-7471B	PPB		04 Jan 2019 16:37:19	U	SMPL
CCV73 - 1	CCV73	28042	5.1698	-7471B	PPB		04 Jan 2019 16:39:37	U	SMPL
CCB73 - 1	CCB73	116	-0.0169	-7471B	PPB		04 Jan 2019 16:41:55	U	SMPL