



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100170

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.3	01/04/2019	14:05
2	K1017-26		1	8	>100.0	>212.0	>212.0	01/04/2019	14:13
3	K1018-01		1	6	>100.0	>212.0	>212.0	01/04/2019	14:20
4	K1018-02		1	7	>100.0	>212.0	>212.0	01/04/2019	14:26
5	K1023-01		1	9	>100.0	>212.0	>212.0	01/04/2019	14:33
6	K1042-02		1	8	>100.0	>212.0	>212.0	01/04/2019	14:40
7	K1043-02		1	6	>100.0	>212.0	>212.0	01/04/2019	14:47
8	K1043-02DUP		1	6	>100.0	>212.0	>212.0	01/04/2019	14:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100170

WorkList Name : FLASHPOINT

WorkList ID : 121028

Date : 1/4/2019 1:53:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/07/2019	Water	K1017-26	Flash Point	Cool 4 deg C	PSEG01	E61	TP-5	01/02/2019	1010A
01/15/2019	Water	K1018-01	Flash Point	Cool 4 deg C	TULL02	D61	TEST-PIT-13	01/02/2019	1010A
01/15/2019	Water	K1018-02	Flash Point	Cool 4 deg C	TULL02	D61	TEST-PIT-D11-12	01/02/2019	1010A
01/15/2019	Water	K1023-01	Flash Point	Cool 4 deg C	SEAC01	D62	PURGE-WATER	01/02/2019	1010A
01/08/2019	Water	K1042-02	Flash Point	Cool 4 deg C	PSEG01	E61	JC-1319-WS	01/03/2019	1010A
01/08/2019	Water	K1043-02	Flash Point	Cool 4 deg C	PSEG01	E61	HB-1319-WS	01/03/2019	1010A

Date/Time 01-04-18 13:55

Received by: AC

Relinquished by: CP

Date/Time 01-04-18 15:10

Received by: CP

Relinquished by: AC