



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB100176

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/04/2019	18:00
2	CAL2	1	Water	NA	NA	21.2	7.01	01/04/2019	18:01
3	CAL3	1	Water	NA	NA	21.2	10.03	01/04/2019	18:02
4	ICV	1	Water	NA	NA	21.2	7.00	01/04/2019	18:03
5	CCV1	1	Water	NA	NA	21.1	2.02	01/04/2019	18:21
6	K1053-02	1	Water	NA	NA	10.6	5.64	01/04/2019	18:22
7	K1053-02DUP	1	Water	NA	NA	10.5	5.65	01/04/2019	18:23
8	K1053-05	1	Water	NA	NA	12.4	5.27	01/04/2019	18:25
9	K1053-08	1	Water	NA	NA	13.5	5.89	01/04/2019	18:26
10	K1053-11	1	Water	NA	NA	13.3	6.11	01/04/2019	18:27
11	K1059-04	1	Water	NA	NA	15.2	6.36	01/04/2019	18:30
12	K1061-02	1	Water	NA	NA	14.7	6.47	01/04/2019	18:31
13	K1064-05	1	Water	NA	NA	18.3	7.04	01/04/2019	18:32
14	K1064-06	1	Water	NA	NA	18.3	7.04	01/04/2019	18:32
15	CCV2	1	Water	NA	NA	21.2	12.02	01/04/2019	18:34

SP 100176 1B100176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion K1053

WorkList ID : 121034

Date : 1/4/2019 5:15:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/04/2019	Water	K1053-02	Corrosivity	Cool 4 deg C	PSEG01	E61	COMP-1-DRUM	01/04/2019	9040C
01/04/2019	Water	K1053-05	Corrosivity	Cool 4 deg C	PSEG01	E61	COMP-2-DRUM	01/04/2019	9040C
01/04/2019	Water	K1053-08	Corrosivity	Cool 4 deg C	PSEG01	E61	COMP-3-DRUM	01/04/2019	9040C
01/04/2019	Water	K1053-11	Corrosivity	Cool 4 deg C	PSEG01	E61	COMP-4-DRUM	01/04/2019	9040C
01/04/2019	Water	K1059-04	Corrosivity	Cool 4 deg C	PSEG01	E61	BUR-COMP-1	01/04/2019	9040C
01/04/2019	Water	K1061-02	Corrosivity	Cool 4 deg C	PSEG01	E61	TRE-COMP-1	01/04/2019	9040C
01/04/2019	Water	K1064-05	Corrosivity	Cool 4 deg C	PSEG01	E61	P-1419-W	01/04/2019	9040C
01/04/2019	Water	K1064-06	Corrosivity	Cool 4 deg C	PSEG01	E61	P-1419-W	01/04/2019	9040C

Date/Time 01-04-19 5:20 PM
 Received by: SP
 Relinquished by: [Signature]

Date/Time 01-04-19 6:30 PM
 Received by: [Signature]
 Relinquished by: SP