

LB100178


ISM02.4

INSTRUMENT ID: CV1

Linear

μ Abs.:

46392



A= 0.0000e+000

B= 2.1851e-004

C= -2.3726e-002

Rho= 0.9996144

Accept=Accepted

*slope
y=1015x-0.02*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	-0.004	-0.004	89	0.000	89					0
0.05	0.050	0.050	0.000	339	0.0 %	339					0
0.2	0.200	0.200	0.000	1026	0.0 %	1026					0
2.5	2.500	2.534	0.034	11704	0.0 %	11704					1
5.0	5.000	5.086	0.086	23385	0.0 %	23385					2
7.5	7.500	7.270	-0.230	33379	0.0 %	33379					-3
10.0	10.000	10.114	0.114	46392	0.0 %	46392					1

*AS
1/4/18*

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	89	-		0 ISM02.4	PPB	1/4/2019 19:39	Std
	0.05 S0.05	339	-		0.05 ISM02.4	PPB	1/4/2019 19:41	Std
	0.2 S0.2	1026	-		0.2 ISM02.4	PPB	1/4/2019 19:43	Std
	2.5 S2.5	11704	-		2.5 ISM02.4	PPB	1/4/2019 19:45	Std
	5 S5	23385	-		5 ISM02.4	PPB	1/4/2019 19:48	Std
	7.5 S7.5	33379	-		7.5 ISM02.4	PPB	1/4/2019 19:50	Std
	10 S10	46392	-		10 ISM02.4	PPB	1/4/2019 19:55	Std
ICV	ICV	17861	3.8791	-	ISM02.4	PPB	1/4/2019 19:58	SMPL
ICB	ICB	40	-0.015	-	ISM02.4	PPB	1/4/2019 20:00	SMPL
CCV078	CCV078	22986	4.999	-	ISM02.4	PPB	1/4/2019 20:02	SMPL
CCB078	CCB078	39	-0.0152	-	ISM02.4	PPB	1/4/2019 20:04	SMPL
PB116166BL	PBS166	117	0.0018	-	ISM02.4	PPB	1/4/2019 20:09	SMPL
J6442-01	MBE7Z7	704507	HIGH	-	ISM02.4	PPB	1/4/2019 20:11	
J6442-02	MBF025	425039	HIGH	-	ISM02.4	PPB	1/4/2019 20:13	
J6442-03	MBE7Z8	191876	41.9036	-	ISM02.4	PPB	1/4/2019 20:15	SMPL
J6442-04	MBE802	763786	HIGH	-	ISM02.4	PPB	1/4/2019 20:18	
J6442-05	MBF023	1276209	HIGH	-	ISM02.4	PPB	1/4/2019 20:19	
J6442-06	MBF023D	1260561	HIGH	-	ISM02.4	PPB	1/4/2019 20:21	
J6442-07	MBF023S	1258958	HIGH	-	ISM02.4	PPB	1/4/2019 20:24	
J6442-01DLX10	MBE7Z7	131985	28.8167	-	ISM02.4	PPB	1/4/2019 20:28	SMPL
J6442-02DLX10	MBF025	60310	13.1548	-	ISM02.4	PPB	1/4/2019 20:31	SMPL
J6442-03DLX10	MBE7Z8	18329	3.9814	-	ISM02.4	PPB	1/4/2019 20:33	SMPL
J6442-04DLX10	MBE902	154349	33.7035	-	ISM02.4	PPB	1/4/2019 20:36	SMPL
CCV079	CCV079	21548	4.6848	-	ISM02.4	PPB	1/4/2019 20:44	SMPL
CCB079	CCB079	-131	-0.0524	-	ISM02.4	PPB	1/4/2019 20:51	SMPL
J6442-01DL2X50		28518	6.2078	-	ISM02.4	PPB	1/4/2019 20:56	SMPL
J6442-02DL2X20		32744	7.1313	-	ISM02.4	PPB	1/4/2019 20:59	SMPL
J6442-04DL2X50		32751	7.1328	-	ISM02.4	PPB	1/4/2019 21:01	SMPL
J6442-05DLX50		259970	56.783	-	ISM02.4	PPB	1/4/2019 21:03	SMPL
J6442-06DLX50		253550	55.3802	-	ISM02.4	PPB	1/4/2019 21:05	SMPL
J6442-07DLX50		254388	55.5633	-	ISM02.4	PPB	1/4/2019 21:08	SMPL
CCV080	CCV080	22504	4.8937	-	ISM02.4	PPB	1/4/2019 21:10	SMPL
CCB080	CCB080	-10	-0.0259	-	ISM02.4	PPB	1/4/2019 21:13	SMPL
J6442-05DL2X500		27927	6.0787	-	ISM02.4	PPB	1/4/2019 21:15	SMPL
J6442-06DL2X500		27839	6.0594	-	ISM02.4	PPB	1/4/2019 21:17	SMPL
J6442-07DL2X500		27282	5.9377	-	ISM02.4	PPB	1/4/2019 21:19	SMPL
CCV081	CCV081	23308	5.0694	-	ISM02.4	PPB	1/4/2019 21:24	SMPL
CCB081	CCB081	214	0.023	-	ISM02.4	PPB	1/4/2019 21:26	SMPL