



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100188

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/07/2019	10:30
2	CAL2	1	Water	NA	NA	21.3	7.01	01/07/2019	10:31
3	CAL3	1	Water	NA	NA	21.3	10.02	01/07/2019	10:32
4	ICV	1	Water	NA	NA	21.3	7.01	01/07/2019	10:33
5	CCV1	1	Water	NA	NA	21.32	2.01	01/07/2019	10:35
6	K1065-01	1	Solid	20.03	20	20.1	8.14	01/07/2019	10:36
7	K1065-01DUP	1	Solid	20.04	20	20.2	8.15	01/07/2019	10:37
8	CCV2	1	Water	NA	NA	21.2	12.02	01/07/2019	10:39

LB 100188

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K1065

WorkList ID : 121072

Date : 1/7/2019 10:22:55 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/04/2019	Solid	K1065-01-17	pH	Cool 4 deg C	TULL01	E51	STOCKPILE	01/04/2019	9045D

Date/Time 01-07-19 10:25 AM
Received by: Jm
Relinquished by: Jm

Date/Time 01-07-19 11:10 AM
Received by: Jm
Relinquished by: Jm