


7470A

LB100189
INSTRUMENT ID: CV2

Linear

 μ Abs.:

39996



A= 0.0000e+000

B= 2.4906e-004

C= -2.4704e-002

Rho= 0.9997030

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	-12	0.000	-12				
0.2	0.200	0.171	-0.029	784	0.0 %	784				
2.5	2.500	2.463	-0.037	9988	0.0 %	9988				
5.0	5.000	5.199	0.199	20975	0.0 %	20975				
7.5	7.500	7.458	-0.042	30043	0.0 %	30043				
10.0	10.000	9.937	-0.063	39996	0.0 %	39996	0			

LB100189 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jan 2019 11:37:49

Sample ID	Extended ID	µ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7470A	PPB	07 Jan 2019 11:45:18	S	Std
0.2 - 1	50.2	784	-	0.2000	7470A	PPB	07 Jan 2019 11:47:27	S	Std
2.5 - 1	52.5	9988	-	2.5000	7470A	PPB	07 Jan 2019 11:49:35	S	Std
5.0 - 1	55	20975	-	5.0000	7470A	PPB	07 Jan 2019 11:51:44	S	Std
7.5 - 1	57.5	30043	-	7.5000	7470A	PPB	07 Jan 2019 11:53:52	S	Std
10.0 - 1	510	39996	-	10.0000	7470A	PPB	07 Jan 2019 11:55:59	S	Std
ICV52 - 1	ICV52	15891	3.9332	-7470A	PPB	PPB	07 Jan 2019 11:59:22	U	SMPL
ICB52 - 1	ICB52	-45	-0.0359	-7470A	PPB	PPB	07 Jan 2019 12:01:28	U	SMPL
CCV82 - 1	CCV82	20303	5.0320	-7470A	PPB	PPB	07 Jan 2019 12:03:35	U	SMPL
CCB82 - 1	CCB82	-64	-0.0406	-7470A	PPB	PPB	07 Jan 2019 12:05:42	U	SMPL
CRA - 1	CRA	767	0.1663	-7470A	PPB	PPB	07 Jan 2019 12:07:49	U	SMPL
HighStd - 1	HighStd	40218	9.9921	-7470A	PPB	PPB	07 Jan 2019 12:09:56	U	SMPL
ChkStd - 1	ChkStd	28066	6.9655	-7470A	PPB	PPB	07 Jan 2019 12:12:02	U	SMPL
PB116175BL - 1	PBW	-85	-0.0459	-7470A	PPB	PPB	07 Jan 2019 12:14:11	U	SMPL
PB116175BS - 1	LCSW	14603	3.6124	-7470A	PPB	PPB	07 Jan 2019 12:16:19	U	SMPL
K1017-26 - 1	TP-5	58	-0.0103	-7470A	PPB	PPB	07 Jan 2019 12:18:27	U	SMPL
K1017-26DUP - 1	TP-5DUP	6	-0.0232	-7470A	PPB	PPB	07 Jan 2019 12:20:36	U	SMPL
K1017-26MS - 1	TP-5MS	14143	3.4978	-7470A	PPB	PPB	07 Jan 2019 12:22:44	U	SMPL
K1017-26MSD - 1	TP-5MSD	13866	3.4288	-7470A	PPB	PPB	07 Jan 2019 12:24:52	U	SMPL
K1042-02 - 1	JC-1319-WS	-90	-0.0471	-7470A	PPB	PPB	07 Jan 2019 12:26:59	U	SMPL
CCV83 - 1	CCV83	19942	4.9421	-7470A	PPB	PPB	07 Jan 2019 12:29:06	U	SMPL
CCB83 - 1	CCB83	-62	-0.0401	-7470A	PPB	PPB	07 Jan 2019 12:31:12	U	SMPL
K1043-02 - 1	HB-1319-WS	35	-0.0160	-7470A	PPB	PPB	07 Jan 2019 12:33:20	U	SMPL
PB116176BL - 1	PBW	-32	-0.0327	-7470A	PPB	PPB	07 Jan 2019 12:35:26	U	SMPL
PB116176BS - 1	LCSW	15694	3.8841	-7470A	PPB	PPB	07 Jan 2019 12:37:33	U	SMPL
K1033-01 - 1	NB-1319-OD	-70	-0.0421	-7470A	PPB	PPB	07 Jan 2019 12:39:40	U	SMPL
K1034-01 - 1	CONTROL-HOUSE-WALL-1-1	2	-0.0242	-7470A	PPB	PPB	07 Jan 2019 12:41:48	U	SMPL
K1034-02 - 1	CONTROL-HOUSE-WALL-1-2	139	0.0099	-7470A	PPB	PPB	07 Jan 2019 12:43:56	U	SMPL
K1034-03 - 1	CONTROL-HOUSE-WALL-1-3	1	-0.0245	-7470A	PPB	PPB	07 Jan 2019 12:46:04	U	SMPL
K1034-04 - 1	CONTROL-HOUSE-WALL-1-4	158	0.0146	-7470A	PPB	PPB	07 Jan 2019 12:48:13	U	SMPL
K1034-05 - 1	CONTROL-HOUSE-WALL-1-5	51	-0.0120	-7470A	PPB	PPB	07 Jan 2019 12:50:22	U	SMPL
K1034-06 - 1	CONTROL-HOUSE-WALL-1-6	-35	-0.0334	-7470A	PPB	PPB	07 Jan 2019 12:52:29	U	SMPL
CCV84 - 1	CCV84	19660	4.8719	-7470A	PPB	PPB	07 Jan 2019 12:54:36	U	SMPL
CCB84 - 1	CCB84	-62	-0.0401	-7470A	PPB	PPB	07 Jan 2019 12:56:44	U	SMPL
K1036-01 - 1	13-A	-15	-0.0284	-7470A	PPB	PPB	07 Jan 2019 12:58:51	U	SMPL
K1036-02 - 1	13-B	-23	-0.0304	-7470A	PPB	PPB	07 Jan 2019 13:00:58	U	SMPL
K1036-03 - 1	13-C	-20	-0.0297	-7470A	PPB	PPB	07 Jan 2019 13:03:06	U	SMPL
K1036-04 - 1	13-D	-41	-0.0349	-7470A	PPB	PPB	07 Jan 2019 13:05:13	U	SMPL
K1036-05 - 1	13-E	-48	-0.0367	-7470A	PPB	PPB	07 Jan 2019 13:07:20	U	SMPL
K1036-06 - 1	13-F	-15	-0.0284	-7470A	PPB	PPB	07 Jan 2019 13:09:28	U	SMPL
K1036-07 - 1	13-G	-7	-0.0264	-7470A	PPB	PPB	07 Jan 2019 13:11:36	U	SMPL
K1036-08 - 1	13-H	1	-0.0245	-7470A	PPB	PPB	07 Jan 2019 13:13:45	U	SMPL
K1036-09 - 1	13-I	0	-0.0247	-7470A	PPB	PPB	07 Jan 2019 13:15:54	U	SMPL
K1036-10 - 1	13-L	-12	-0.0277	-7470A	PPB	PPB	07 Jan 2019 13:18:01	U	SMPL
CCV85 - 1	CCV85	19700	4.8819	-7470A	PPB	PPB	07 Jan 2019 13:20:09	U	SMPL
CCB85 - 1	CCB85	-48	-0.0367	-7470A	PPB	PPB	07 Jan 2019 13:22:17	U	SMPL
K1036-11 - 1	13-M	0	-0.0247	-7470A	PPB	PPB	07 Jan 2019 13:24:24	U	SMPL
K1040-03 - 1	RBR-200030-RBR-200028-RT-3483	-9	-0.0269	-7470A	PPB	PPB	07 Jan 2019 13:26:32	U	SMPL
K1041-04 - 1	OS-1-5-COMP	-22	-0.0302	-7470A	PPB	PPB	07 Jan 2019 13:28:39	U	SMPL
K1041-04DUP - 1	OS-1-5-COMPDUP	-22	-0.0302	-7470A	PPB	PPB	07 Jan 2019 13:30:46	U	SMPL
K1041-04MS - 1	OS-1-5-COMPMS	14361	3.5521	-7470A	PPB	PPB	07 Jan 2019 13:32:53	U	SMPL
K1041-04MSD - 1	OS-1-5-COMPMSD	14292	3.5349	-7470A	PPB	PPB	07 Jan 2019 13:35:01	U	SMPL
PB116177BL - 1	PBW	-75	-0.0434	-7470A	PPB	PPB	07 Jan 2019 13:37:10	U	SMPL
PB116177BS - 1	LCSW	15596	3.8597	-7470A	PPB	PPB	07 Jan 2019 13:39:18	U	SMPL
K1036-12 - 1	13-N	-61	-0.0399	-7470A	PPB	PPB	07 Jan 2019 13:41:27	U	SMPL
K1036-12DUP - 1	13-NDUP	-4	-0.0257	-7470A	PPB	PPB	07 Jan 2019 13:43:34	U	SMPL
CCV86 - 1	CCV86	20578	5.1005	-7470A	PPB	PPB	07 Jan 2019 13:45:41	U	SMPL
CCB86 - 1	CCB86	-66	-0.0411	-7470A	PPB	PPB	07 Jan 2019 13:47:49	U	SMPL
K1036-12MS - 1	13-NMS	14049	3.4744	-7470A	PPB	PPB	07 Jan 2019 13:49:57	U	SMPL
K1036-12MSD - 1	13-NMSD	13798	3.4119	-7470A	PPB	PPB	07 Jan 2019 13:52:04	U	SMPL
K1036-13 - 1	13-O	-48	-0.0367	-7470A	PPB	PPB	07 Jan 2019 13:54:11	U	SMPL
K1036-14 - 1	13-P	-26	-0.0312	-7470A	PPB	PPB	07 Jan 2019 13:56:19	U	SMPL
K1036-15 - 1	13-Q	-22	-0.0302	-7470A	PPB	PPB	07 Jan 2019 13:58:26	U	SMPL
K1036-16 - 1	13-R	-9	-0.0269	-7470A	PPB	PPB	07 Jan 2019 14:00:34	U	SMPL
K1036-17 - 1	13-S	-10	-0.0272	-7470A	PPB	PPB	07 Jan 2019 14:02:43	U	SMPL
K1036-18 - 1	13-T	8	-0.0227	-7470A	PPB	PPB	07 Jan 2019 14:04:51	U	SMPL
K1036-19 - 1	13-U	6	-0.0232	-7470A	PPB	PPB	07 Jan 2019 14:07:00	U	SMPL
K1036-20 - 1	13-V	8	-0.0227	-7470A	PPB	PPB	07 Jan 2019 14:09:07	U	SMPL
CCV87 - 1	CCV87	20953	5.1939	-7470A	PPB	PPB	07 Jan 2019 14:11:16	U	SMPL
CCB87 - 1	CCB87	-69	-0.0419	-7470A	PPB	PPB	07 Jan 2019 14:13:23	U	SMPL
K1036-21 - 1	13-W	-10	-0.0272	-7470A	PPB	PPB	07 Jan 2019 14:15:31	U	SMPL
K1036-22 - 1	13-J	4	-0.0237	-7470A	PPB	PPB	07 Jan 2019 14:17:39	U	SMPL
PB116187BL - 1	PBW	-14	-0.0282	-7470A	PPB	PPB	07 Jan 2019 14:19:46	U	SMPL
PB116187BS - 1	LCSW	15123	3.7419	-7470A	PPB	PPB	07 Jan 2019 14:21:54	U	SMPL
K1042-01 - 1	JC-1319-WS	-65	-0.0409	-7470A	PPB	PPB	07 Jan 2019 14:24:02	U	SMPL
K1043-01 - 1	HB-1319-WS	1180	0.2692	-7470A	PPB	PPB	07 Jan 2019 14:26:10	U	SMPL
K1043-01DUP - 1	HB-1319-WS DUP	1096	0.2483	-7470A	PPB	PPB	07 Jan 2019 14:28:19	U	SMPL
K1043-01MS - 1	HB-1319-WSMS	16909	4.1867	-7470A	PPB	PPB	07 Jan 2019 14:30:28	U	SMPL
K1043-01MSD - 1	HB-1319-WSMSD	16719	4.1394	-7470A	PPB	PPB	07 Jan 2019 14:32:37	U	SMPL
CCV88 - 1	CCV88	20595	5.1048	-7470A	PPB	PPB	07 Jan 2019 14:34:44	U	SMPL
CCB88 - 1	CCB88	-90	-0.0471	-7470A	PPB	PPB	07 Jan 2019 14:36:53	U	SMPL
PB116175TB - 1	PB116145TB	27	-0.0180	-7470A	PPB	PPB	07 Jan 2019 14:39:01	U	SMPL
PB116176TB - 1	PB116148TB	9	-0.0225	-7470A	PPB	PPB	07 Jan 2019 14:41:09	U	SMPL
PB116177TB - 1	PB116149TB	19	-0.0200	-7470A	PPB	PPB	07 Jan 2019 14:43:17	U	SMPL
CCV89 - 1	CCV89	20114	4.9850	-7470A	PPB	PPB	07 Jan 2019 14:45:24	U	SMPL
CCB89 - 1	CCB89	-50	-0.0372	-7470A	PPB	PPB	07 Jan 2019 14:47:32	U	SMPL
K1017-26LX5 - 1		29	-0.0175	-7470A	PPB	PPB	07 Jan 2019 14:49:41	U	SMPL
K1017-26A - 1		16505	4.0861	-7470A	PPB	PPB	07 Jan 2019 14:51:49	U	SMPL
K1041-04LX5 - 1		-60	-0.0396	-7470A	PPB	PPB	07 Jan 2019 14:53:57	U	SMPL
K1041-04A - 1		16066	3.9768	-7470A	PPB	PPB	07 Jan 2019 14:56:06	U	SMPL
K1036-12LX5 - 1		16	-0.0207	-7470A	PPB	PPB	07 Jan 2019 15:01:13	U	SMPL
K1036-12A - 1		16350	4.0475	-7470A	PPB	PPB	07 Jan 2019 15:03:21	U	SMPL
K1043-01LX5 - 1		448	0.0869	-7470A	PPB	PPB	07 Jan 2019 15:05:29	U	SMPL

LB100189 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jan 2019 11:37:49

Sample ID	Extended ID	μ Abs	Conc	Stnd Conc	Method	Units	Date	Type	Type
K1043-01A - 1		18398	4.5576		-7470A	PPB	07 Jan 2019 15:07:36	U	SMPL
CCV90 - 1	CCV90	19902	4.9322		-7470A	PPB	07 Jan 2019 15:09:45	U	SMPL
CCB90 - 1	CCB90	-80	-0.0446		-7470A	PPB	07 Jan 2019 15:11:53	U	SMPL