



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/8/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 17:00  
In Date: 01/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:10  
Out Date: 01/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1

QC:LB100197

| Lab ID   | Client Sample ID      | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|-----------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1071-01 | CPSB-17(10-15)        | 1     | 1.13              | 9.73                          | 9.04                              | 92.0    |
| K1071-02 | CPSB-17(20-25)        | 2     | 1.15              | 9.55                          | 7.29                              | 73.1    |
| K1071-03 | CPSB-16(10-15)        | 3     | 1.16              | 9.85                          | 9.11                              | 91.5    |
| K1071-04 | CPSB-16(20-25)        | 4     | 1.18              | 9.78                          | 3.17                              | 23.1    |
| K1071-05 | CPSB-15(10-15)        | 5     | 1.18              | 9.94                          | 8.01                              | 78.0    |
| K1071-06 | CPSB-15(20-25)        | 6     | 1.14              | 9.66                          | 4.18                              | 35.7    |
| K1071-07 | CPSB-1(10-15)         | 7     | 1.11              | 9.94                          | 8.12                              | 79.4    |
| K1071-08 | CPSB-1(20-25)         | 8     | 1.16              | 9.68                          | 4.4                               | 38.0    |
| K1071-09 | CPSB-3(10-15)         | 9     | 1.11              | 9.64                          | 8.53                              | 87.0    |
| K1071-10 | CPSB-3(20-25)         | 10    | 1.13              | 9.65                          | 8.41                              | 85.4    |
| K1072-01 | FEEDAR-ROW-SLABS      | 11    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1072-02 | CAPACITOR-SLABS       | 12    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1072-03 | TRANSFORMER-PIERS     | 13    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1072-04 | SWITCH-PIERS          | 14    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1072-05 | 26KV-POTHEAD-SLABS    | 15    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1073-01 | RICH-REC-ROAD-STONE-9 | 16    | 1.13              | 9.73                          | 9.27                              | 94.7    |
| K1073-02 | RICH-REC-ROAD-STONE-9 | 17    | 1.19              | 9.8                           | 9.4                               | 95.4    |
| K1074-01 | W12-1                 | 18    | 1.15              | 9.95                          | 8.1                               | 79.0    |
| K1074-02 | W12-1                 | 19    | 1.14              | 9.93                          | 8.06                              | 78.7    |
| K1074-03 | W12-2                 | 20    | 1.13              | 9.58                          | 7.31                              | 73.1    |
| K1074-04 | W12-2                 | 21    | 1.12              | 9.65                          | 7.5                               | 74.8    |
| K1074-05 | W12-3                 | 22    | 1.1               | 9.84                          | 7.7                               | 75.5    |
| K1074-06 | W12-3                 | 23    | 1.17              | 9.69                          | 7.29                              | 71.8    |
| K1074-07 | W12-4                 | 24    | 1.19              | 9.77                          | 7.6                               | 74.7    |
| K1074-08 | W12-4                 | 25    | 1.18              | 9.98                          | 7.3                               | 69.5    |
| K1074-09 | W12-5                 | 26    | 1.17              | 9.97                          | 8.33                              | 81.4    |
| K1074-10 | W12-5                 | 27    | 1.1               | 9.85                          | 8.00                              | 78.9    |
| K1074-11 | W12-6                 | 28    | 1.1               | 9.97                          | 7.52                              | 72.4    |
| K1074-12 | W12-6                 | 29    | 1.18              | 9.79                          | 7.11                              | 68.9    |
| K1074-13 | W12-7                 | 30    | 1.15              | 9.95                          | 8.17                              | 79.8    |
| K1074-14 | W12-7                 | 31    | 1.12              | 9.64                          | 7.97                              | 80.4    |
| K1074-15 | W12-8                 | 32    | 1.13              | 9.57                          | 8.39                              | 86.0    |
| K1074-16 | W12-8                 | 33    | 1.18              | 9.72                          | 8.4                               | 84.5    |
| K1074-17 | W12-9                 | 34    | 1.13              | 9.61                          | 8.29                              | 84.4    |
| K1074-18 | W12-9                 | 35    | 1.15              | 9.74                          | 8.69                              | 87.8    |
| K1074-19 | W12-10                | 36    | 1.19              | 9.97                          | 8.55                              | 83.8    |
| K1074-20 | W12-10                | 37    | 1.17              | 9.98                          | 8.22                              | 80.0    |
| K1075-01 | BASIN-1               | 38    | 1.1               | 9.54                          | 7.72                              | 78.4    |
| K1075-02 | BASIN-1               | 39    | 1.18              | 9.97                          | 8.15                              | 79.3    |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/8/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 17:00  
In Date: 01/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:10  
Out Date: 01/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1

QC:LB100197

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1075-03 | BASIN-2          | 40    | 1.16              | 9.51                          | 7.69                              | 78.2    |
| K1075-04 | BASIN-2          | 41    | 1.17              | 9.68                          | 7.47                              | 74.0    |
| K1075-05 | BASIN-3          | 42    | 1.13              | 9.55                          | 8.16                              | 83.5    |
| K1075-06 | BASIN-3          | 43    | 1.18              | 9.56                          | 8.45                              | 86.8    |
| K1075-07 | BASIN-4          | 44    | 1.16              | 9.79                          | 7.21                              | 70.1    |
| K1075-08 | BASIN-4          | 45    | 1.18              | 9.97                          | 7.9                               | 76.5    |
| K1075-09 | BASIN-5          | 46    | 1.18              | 9.63                          | 8.57                              | 87.5    |
| K1075-10 | BASIN-5          | 47    | 1.18              | 9.95                          | 7.87                              | 76.3    |
| K1075-11 | BASIN-6          | 48    | 1.17              | 9.59                          | 7.31                              | 72.9    |
| K1075-12 | BASIN-6          | 49    | 1.19              | 9.7                           | 7.25                              | 71.2    |
| K1075-13 | BASIN-7          | 50    | 1.1               | 9.95                          | 7.6                               | 73.4    |
| K1075-14 | BASIN-7          | 51    | 1.15              | 9.86                          | 7.67                              | 74.9    |
| K1075-15 | BASIN-8          | 52    | 1.18              | 9.94                          | 8.08                              | 78.8    |
| K1075-16 | BASIN-8          | 53    | 1.1               | 9.55                          | 7.48                              | 75.5    |
| K1075-17 | BASIN-9          | 54    | 1.17              | 9.86                          | 7.82                              | 76.5    |
| K1075-18 | BASIN-9          | 55    | 1.17              | 9.68                          | 7.73                              | 77.1    |
| K1075-19 | BASIN-10         | 56    | 1.13              | 9.74                          | 8.33                              | 83.6    |
| K1075-20 | BASIN-10         | 57    | 1.12              | 9.92                          | 8.24                              | 80.9    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$