

LB100199


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

39979



A= 0.0000e+000

B= 2.5264e-004

C= 8.3645e-003

Rho= 0.9997161

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	-14	0.000	-14				
0.2	0.200	0.189	-0.011	715	0.0 %	715				
2.5	2.500	2.582	0.082	10187	0.0 %	10187				
5.0	5.000	4.981	-0.019	19682	0.0 %	19682				
7.5	7.500	7.335	-0.165	28999	0.0 %	28999				
10.0	10.000	10.109	0.109	39979	0.0 %	39979	0			

LB100199 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jan 2019 09:52:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-14	-	0.0000	7470A	PPB	08 Jan 2019 10:03:01	S	Std
0.2 - 1	50.2	715	-	0.2000	7470A	PPB	08 Jan 2019 10:05:10	S	Std
2.5 - 1	50.5	10187	-	2.5000	7470A	PPB	08 Jan 2019 10:07:19	S	Std
5.0 - 1	55.0	19682	-	5.0000	7470A	PPB	08 Jan 2019 10:09:28	S	Std
7.5 - 1	57.5	28999	-	7.5000	7470A	PPB	08 Jan 2019 10:11:36	S	Std
10.0 - 1	510.0	39979	-	10.0000	7470A	PPB	08 Jan 2019 10:13:43	S	Std
ICV53 - 1	ICV53	15044	3.8091	-	7470A	PPB	08 Jan 2019 10:20:24	U	SMPL
ICB53 - 1	ICB53	-51	-0.0045	-	7470A	PPB	08 Jan 2019 10:22:30	U	SMPL
CCV91 - 1	CCV91	19447	4.9215	-	7470A	PPB	08 Jan 2019 10:24:37	U	SMPL
CCB91 - 1	CCB91	-73	-0.0101	-	7470A	PPB	08 Jan 2019 10:26:44	U	SMPL
CRA - 1	CRA	687	0.1819	-	7470A	PPB	08 Jan 2019 10:28:51	U	SMPL
HighStd - 1	HighStd	40497	10.2395	-	7470A	PPB	08 Jan 2019 10:30:58	U	SMPL
ChkStd - 1	ChkStd	27641	6.9916	-	7470A	PPB	08 Jan 2019 10:33:05	U	SMPL
PB116218BL - 1		-85	-0.0131	-	7470A	PPB	08 Jan 2019 10:35:14	U	SMPL
PB116218BS - 1	LCSW	14920	3.7778	-	7470A	PPB	08 Jan 2019 10:37:22	U	SMPL
K1010-02 - 1	OK-03-010419	-51	-0.0045	-	7470A	PPB	08 Jan 2019 10:39:30	U	SMPL
K1010-02DUP - 1	OK-03-010419DUP	35	0.0172	-	7470A	PPB	08 Jan 2019 10:41:38	U	SMPL
K1010-02MS - 1	OK-03-010419MS	14235	3.6047	-	7470A	PPB	08 Jan 2019 10:43:47	U	SMPL
K1010-02MSD - 1	OK-03-010419MSD	13767	3.4865	-	7470A	PPB	08 Jan 2019 10:45:55	U	SMPL
K1012-02 - 1	EO-02-010419-A	-69	-0.0091	-	7470A	PPB	08 Jan 2019 10:48:02	U	SMPL
CCV92 - 1	CCV92	19248	4.8712	-	7470A	PPB	08 Jan 2019 10:50:09	U	SMPL
CCB92 - 1	CCB92	-89	-0.0141	-	7470A	PPB	08 Jan 2019 10:52:16	U	SMPL
K1012-04 - 1	EO-02-010419-B	-27	0.0015	-	7470A	PPB	08 Jan 2019 10:54:23	U	SMPL
K1012-06 - 1	EO-02-010419-C	-24	0.0023	-	7470A	PPB	08 Jan 2019 10:56:30	U	SMPL
K1062-02 - 1	TP-(1-10)-COMP	11	0.0111	-	7470A	PPB	08 Jan 2019 10:58:37	U	SMPL
K1064-03 - 1	P-1419-S	17	0.0127	-	7470A	PPB	08 Jan 2019 11:03:03	U	SMPL
K1066-03 - 1	TN-1419	8	0.0104	-	7470A	PPB	08 Jan 2019 11:05:10	U	SMPL
PB116219BL - 1	PBW	-15	0.0046	-	7470A	PPB	08 Jan 2019 11:07:18	U	SMPL
PB116219BS - 1	LCSW	16024	4.0567	-	7470A	PPB	08 Jan 2019 11:09:26	U	SMPL
K1053-02 - 1	COMP-1-DRUM	-10	0.0058	-	7470A	PPB	08 Jan 2019 11:11:35	U	SMPL
K1053-02DUP - 1	COMP-1-DRUMDUP	40	0.0185	-	7470A	PPB	08 Jan 2019 11:13:43	U	SMPL
K1053-02MS - 1	COMP-1-DRUMMS	13471	3.4117	-	7470A	PPB	08 Jan 2019 11:15:50	U	SMPL
CCV93 - 1	CCV93	18995	4.8073	-	7470A	PPB	08 Jan 2019 11:17:58	U	SMPL
CCB93 - 1	CCB93	-49	-0.0040	-	7470A	PPB	08 Jan 2019 11:20:05	U	SMPL
K1053-02MSD - 1	COMP-1-DRUMMSD	14274	3.6145	-	7470A	PPB	08 Jan 2019 11:22:12	U	SMPL
K1053-05 - 1	COMP-2-DRUM	-16	0.0043	-	7470A	PPB	08 Jan 2019 11:24:19	U	SMPL
K1053-08 - 1	COMP-3-DRUM	5	0.0096	-	7470A	PPB	08 Jan 2019 11:26:26	U	SMPL
K1053-11 - 1	COMP-4-DRUM	0	0.0084	-	7470A	PPB	08 Jan 2019 11:28:33	U	SMPL
K1059-04 - 1	BUR-COMP-1	118	0.0382	-	7470A	PPB	08 Jan 2019 11:30:41	U	SMPL
K1061-02 - 1	TRE-COMP-1	103	0.0344	-	7470A	PPB	08 Jan 2019 11:32:49	U	SMPL
K1064-06 - 1	P-1419-W	89	0.0308	-	7470A	PPB	08 Jan 2019 11:34:57	U	SMPL
PB116220BL - 1	PBW	37	0.0177	-	7470A	PPB	08 Jan 2019 11:37:06	U	SMPL
PB116220BS - 1	LCSW	16689	4.2247	-	7470A	PPB	08 Jan 2019 11:39:14	U	SMPL
K1053-01 - 1	COMP-1-DRUM	274	0.0776	-	7470A	PPB	08 Jan 2019 11:41:22	U	SMPL
CCV94 - 1	CCV94	19067	4.8255	-	7470A	PPB	08 Jan 2019 11:43:29	U	SMPL
CCB94 - 1	CCB94	-53	-0.0050	-	7470A	PPB	08 Jan 2019 11:45:38	U	SMPL
K1053-01DUP - 1	COMP-1-DRUMDUP	226	0.0655	-	7470A	PPB	08 Jan 2019 11:47:45	U	SMPL
K1053-01MS - 1	COMP-1-DRUMMS	14104	3.5716	-	7470A	PPB	08 Jan 2019 11:49:52	U	SMPL
K1053-01MSD - 1	COMP-1-DRUMMSD	14872	3.7656	-	7470A	PPB	08 Jan 2019 11:52:00	U	SMPL
K1053-04 - 1	COMP-2-DRUM	3	0.0091	-	7470A	PPB	08 Jan 2019 11:54:08	U	SMPL
K1053-07 - 1	COMP-3-DRUM	-25	0.0020	-	7470A	PPB	08 Jan 2019 11:56:16	U	SMPL
K1053-10 - 1	COMP-4-DRUM	160	0.0488	-	7470A	PPB	08 Jan 2019 11:58:24	U	SMPL
K1059-03 - 1	BUR-COMP-1	-77	-0.0111	-	7470A	PPB	08 Jan 2019 12:00:32	U	SMPL
K1061-01 - 1	TRE-COMP-1	176	0.0528	-	7470A	PPB	08 Jan 2019 12:02:41	U	SMPL
K1064-05 - 1	P-1419-W	2530	0.6475	-	7470A	PPB	08 Jan 2019 12:04:50	U	SMPL
CCV95 - 1	CCV95	19524	4.9409	-	7470A	PPB	08 Jan 2019 12:06:57	U	SMPL
CCB95 - 1	CCB95	-45	-0.0030	-	7470A	PPB	08 Jan 2019 12:09:05	U	SMPL
PB116218TB - 1	PB116188TB	33	0.0167	-	7470A	PPB	08 Jan 2019 12:11:12	U	SMPL
PB116219TB - 1	PB116190TB	13	0.0116	-	7470A	PPB	08 Jan 2019 12:13:19	U	SMPL
K1010-02LX5 - 1		15	0.0122	-	7470A	PPB	08 Jan 2019 12:15:26	U	SMPL
K1010-02A - 1		15798	3.9996	-	7470A	PPB	08 Jan 2019 12:17:34	U	SMPL
K1053-02LX5 - 1		15	0.0122	-	7470A	PPB	08 Jan 2019 12:19:41	U	SMPL
K1053-02A - 1		15873	4.0185	-	7470A	PPB	08 Jan 2019 12:21:49	U	SMPL
K1053-01LX5 - 1		26	0.0149	-	7470A	PPB	08 Jan 2019 12:23:57	U	SMPL
K1053-01A - 1		16368	4.1436	-	7470A	PPB	08 Jan 2019 12:26:05	U	SMPL
CCV96 - 1	CCV96	19231	4.8669	-	7470A	PPB	08 Jan 2019 12:28:14	U	SMPL
CCB96 - 1	CCB96	-12	0.0053	-	7470A	PPB	08 Jan 2019 12:30:23	U	SMPL