

LB100200


7471B

INSTRUMENT ID-CV1

Linear

 μ Abs.:

54713



A= 0.0000e+000

B= 1.8080e-004

C= -1.1982e-002

Rho= 0.9997557

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	55	0.000	55				
0.2	0.200	0.168	-0.032	996	0.0 %	996				
2.5	2.500	2.446	-0.054	13596	0.0 %	13596				
5.0	5.000	5.103	0.103	28291	0.0 %	28291				
7.5	7.500	7.605	0.105	42129	0.0 %	42129				
10.0	10.000	9.880	-0.120	54713	0.0 %	54713				

LB100200 - INSTRUMENT ID - CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Jan 2019 14:02:15

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	55	-	0.0000	7471B	PPB	08 Jan 2019 14:03:22	S	Std
0.2 - 1	50.2	996	-	0.2000	7471B	PPB	08 Jan 2019 14:05:40	S	Std
2.5 - 1	52.5	13596	-	2.5000	7471B	PPB	08 Jan 2019 14:07:58	S	Std
5.0 - 1	55.0	28291	-	5.0000	7471B	PPB	08 Jan 2019 14:10:16	S	Std
7.5 - 1	57.5	42129	-	7.5000	7471B	PPB	08 Jan 2019 14:12:33	S	Std
10.0 - 1	510.0	54713	-	10.0000	7471B	PPB	08 Jan 2019 14:14:52	S	Std
ICV54 - 1	ICV54	21607	3.8945	-7471B	PPB	PPB	08 Jan 2019 14:17:53	U	SMPL
ICB54 - 1	ICB54	-53	-0.0216	-7471B	PPB	PPB	08 Jan 2019 14:20:08	U	SMPL
CCV97 - 1	CCV97	28255	5.0965	-7471B	PPB	PPB	08 Jan 2019 14:22:25	U	SMPL
CCB97 - 1	CCB97	-76	-0.0257	-7471B	PPB	PPB	08 Jan 2019 14:24:41	U	SMPL
CRA - 1	CRA	958	0.1612	-7471B	PPB	PPB	08 Jan 2019 14:26:58	U	SMPL
HighStd - 1	HighStd	55248	9.9767	-7471B	PPB	PPB	08 Jan 2019 14:29:14	U	SMPL
ChkStd - 1	ChkStd	38436	6.9372	-7471B	PPB	PPB	08 Jan 2019 14:31:30	U	SMPL
PB116231BL - 1	PBS	-92	-0.0286	-7471B	PPB	PPB	08 Jan 2019 14:34:23	U	SMPL
PB116231BS - 1	LCSS	21695	3.9104	-7471B	PPB	PPB	08 Jan 2019 14:36:40	U	SMPL
J6559-11 - 1	P001-SS128-0206-01	27040	4.8768	-7471B	PPB	PPB	08 Jan 2019 14:38:57	U	SMPL
J6559-12 - 1	P001-SS130-0206-01	8310	1.4904	-7471B	PPB	PPB	08 Jan 2019 14:41:15	U	SMPL
J6559-13 - 1	P001-SS131-0206-01	5031	0.8976	-7471B	PPB	PPB	08 Jan 2019 14:43:34	U	SMPL
J6559-14 - 1	P001-SS132-0002-01	22368	4.0321	-7471B	PPB	PPB	08 Jan 2019 14:45:52	U	SMPL
J6559-15 - 1	P001-SS133-0206-01	5003	0.8925	-7471B	PPB	PPB	08 Jan 2019 14:48:08	U	SMPL
CCV98 - 1	CCV98	28228	5.0916	-7471B	PPB	PPB	08 Jan 2019 14:50:26	U	SMPL
CCB98 - 1	CCB98	-112	-0.0322	-7471B	PPB	PPB	08 Jan 2019 14:52:42	U	SMPL
J6559-16 - 1	P001-SS134-0612-01	4534	0.8078	-7471B	PPB	PPB	08 Jan 2019 14:54:58	U	SMPL
J6559-17 - 1	P001-SS136-0002-01	9352	1.6788	-7471B	PPB	PPB	08 Jan 2019 14:57:14	U	SMPL
J6559-18 - 1	P001-SS136-0612-01	4383	0.7805	-7471B	PPB	PPB	08 Jan 2019 14:59:31	U	SMPL
K1010-01 - 1	OK-03-010419	1012	0.1710	-7471B	PPB	PPB	08 Jan 2019 15:01:47	U	SMPL
K1012-01 - 1	EO-02-010419-A	1945	0.3397	-7471B	PPB	PPB	08 Jan 2019 15:04:03	U	SMPL
K1012-01DUP - 1	EO-02-010419-ADUP	1598	0.2769	-7471B	PPB	PPB	08 Jan 2019 15:13:19	U	SMPL
K1012-01MS - 1	EO-02-010419-AMS	23067	4.1585	-7471B	PPB	PPB	08 Jan 2019 15:15:36	U	SMPL
K1012-01MSD - 1	EO-02-010419-AMSD	21266	3.8329	-7471B	PPB	PPB	08 Jan 2019 15:17:54	U	SMPL
K1012-03 - 1	EO-02-010419-B	2457	0.4322	-7471B	PPB	PPB	08 Jan 2019 15:20:12	U	SMPL
K1012-05 - 1	EO-02-010419-C	2152	0.3771	-7471B	PPB	PPB	08 Jan 2019 15:22:29	U	SMPL
CCV99 - 1	CCV99	27799	5.0140	-7471B	PPB	PPB	08 Jan 2019 15:24:46	U	SMPL
CCB99 - 1	CCB99	-93	-0.0288	-7471B	PPB	PPB	08 Jan 2019 15:27:02	U	SMPL
K1052-02 - 1	SSSI-0103-S-DH-8A	3117	0.5516	-7471B	PPB	PPB	08 Jan 2019 15:29:20	U	SMPL
K1058-01 - 1	TP-(1-10)-COMP	25360	4.5730	-7471B	PPB	PPB	08 Jan 2019 15:31:36	U	SMPL
K1062-01 - 1	TP-(1-10)-COMP	21037	3.7915	-7471B	PPB	PPB	08 Jan 2019 15:33:53	U	SMPL
K1064-01 - 1	P-1419-S	17265	3.1095	-7471B	PPB	PPB	08 Jan 2019 15:36:11	U	SMPL
K1065-01 - 1	STOCKPILE	13743	2.4727	-7471B	PPB	PPB	08 Jan 2019 15:38:30	U	SMPL
K1066-01 - 1	TN-1419	4065	0.7230	-7471B	PPB	PPB	08 Jan 2019 15:40:48	U	SMPL
K1073-01 - 1	RICH-REC-ROAD-STONE-9	636	0.1030	-7471B	PPB	PPB	08 Jan 2019 15:43:05	U	SMPL
PB116232BS - 1	LCSS	19758	3.5602	-7471B	PPB	PPB	08 Jan 2019 15:47:41	U	SMPL
K1074-01 - 1	W12-1	1226	0.2097	-7471B	PPB	PPB	08 Jan 2019 15:49:58	U	SMPL
CCV100 - 1 CCV01	CCV100 CCV01	27635	4.9844	-7471B	PPB	PPB	08 Jan 2019 15:52:15	U	SMPL
CCB100 - 1 CCB01	CCB100 CCB01	-67	-0.0241	-7471B	PPB	PPB	08 Jan 2019 15:54:31	U	SMPL
K1074-01DUP - 1	W12-1DUP	1143	0.1947	-7471B	PPB	PPB	08 Jan 2019 16:01:59	U	SMPL
K1074-01MS - 1	W12-1MS	20809	3.7502	-7471B	PPB	PPB	08 Jan 2019 16:04:15	U	SMPL
K1074-01MSD - 1	W12-1MSD	20970	3.7793	-7471B	PPB	PPB	08 Jan 2019 16:06:32	U	SMPL
K1074-03 - 1	W12-2	2011	0.3516	-7471B	PPB	PPB	08 Jan 2019 16:08:49	U	SMPL
K1074-05 - 1	W12-3	1737	0.3021	-7471B	PPB	PPB	08 Jan 2019 16:11:06	U	SMPL
K1074-07 - 1	W12-4	1478	0.2552	-7471B	PPB	PPB	08 Jan 2019 16:13:23	U	SMPL
K1074-09 - 1	W12-5	1081	0.1835	-7471B	PPB	PPB	08 Jan 2019 16:15:40	U	SMPL
K1074-11 - 1	W12-6	918	0.1540	-7471B	PPB	PPB	08 Jan 2019 16:17:58	U	SMPL
K1074-13 - 1	W12-7	827	0.1375	-7471B	PPB	PPB	08 Jan 2019 16:20:17	U	SMPL
K1074-15 - 1	W12-8	2567	0.4521	-7471B	PPB	PPB	08 Jan 2019 16:22:33	U	SMPL
CCV101 - 1 CCV02	CCV101 CCV02	27508	4.9614	-7471B	PPB	PPB	08 Jan 2019 16:24:50	U	SMPL
CCB101 - 1 CCB02	CCB101 CCB02	-66	-0.0239	-7471B	PPB	PPB	08 Jan 2019 16:27:06	U	SMPL
K1074-17 - 1	W12-9	718	0.1178	-7471B	PPB	PPB	08 Jan 2019 16:29:24	U	SMPL
K1074-19 - 1	W12-10	2306	0.4049	-7471B	PPB	PPB	08 Jan 2019 16:31:41	U	SMPL
K1075-01 - 1	BASIN-1	3759	0.6676	-7471B	PPB	PPB	08 Jan 2019 16:33:57	U	SMPL
K1075-03 - 1	BASIN-2	2264	0.3973	-7471B	PPB	PPB	08 Jan 2019 16:36:14	U	SMPL
K1075-05 - 1	BASIN-3	2493	0.4387	-7471B	PPB	PPB	08 Jan 2019 16:38:31	U	SMPL
K1075-07 - 1	BASIN-4	1695	0.2945	-7471B	PPB	PPB	08 Jan 2019 16:40:48	U	SMPL
K1075-09 - 1	BASIN-5	1161	0.1979	-7471B	PPB	PPB	08 Jan 2019 16:43:05	U	SMPL
K1075-11 - 1	BASIN-6	1527	0.2641	-7471B	PPB	PPB	08 Jan 2019 16:45:23	U	SMPL
K1075-13 - 1	BASIN-7	1216	0.2079	-7471B	PPB	PPB	08 Jan 2019 16:47:41	U	SMPL
K1075-15 - 1	BASIN-8	1366	0.2350	-7471B	PPB	PPB	08 Jan 2019 16:49:58	U	SMPL
CCV102 - 1 CCV03	CCV102 CCV03	27455	4.9518	-7471B	PPB	PPB	08 Jan 2019 16:52:15	U	SMPL
CCB102 - 1 CCB03	CCB102 CCB03	-80	-0.0264	-7471B	PPB	PPB	08 Jan 2019 16:54:32	U	SMPL
K1075-17 - 1	BASIN-9	9003	1.6157	-7471B	PPB	PPB	08 Jan 2019 16:56:50	U	SMPL
K1075-19 - 1	BASIN-10	2467	0.4340	-7471B	PPB	PPB	08 Jan 2019 16:59:07	U	SMPL
K1012-01LX5 - 1		357	0.0526	-7471B	PPB	PPB	08 Jan 2019 17:01:24	U	SMPL
K1012-01A - 1		22816	4.1131	-7471B	PPB	PPB	08 Jan 2019 17:03:41	U	SMPL
K1074-01LX5 - 1		158	0.0166	-7471B	PPB	PPB	08 Jan 2019 17:05:58	U	SMPL
K1074-01A - 1		21588	3.8911	-7471B	PPB	PPB	08 Jan 2019 17:18:27	U	SMPL
PB116232BL - 1	PBS	-96	-0.0293	-7471B	PPB	PPB	08 Jan 2019 17:20:44	U	SMPL
CCV103 - 1 CCV04	CCV103 CCV04	26294	4.7419	-7471B	PPB	PPB	08 Jan 2019 17:23:02	U	SMPL
CCB103 - 1 CCB04	CCB103 CCB04	-25	-0.0165	-7471B	PPB	PPB	08 Jan 2019 17:25:20	U	SMPL