

2B/00206

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/8/2019 14:44

Test: ^{Reactive} ~~Total~~ CN

Sample Id	^{mg/L} Result	Dil. 1 +	Response	Errors
ICV1	95.944	0.0	0.087	
ICB1	0.142	0.0	0.004	
CCV1	242.414	0.0	0.214	
CCB1	-0.126	0.0	0.003	
PB116240BL	-0.019	0.0	0.003	
K1062-02	-0.183	0.0	0.003	
K1064-03	-0.055	0.0	0.003	
K1065-02	0.735	0.0	0.004	
K1065-02DUP	0.499	0.0	0.004	
PB116243BL	0.045	0.0	0.004	
K1064-06	11.505	0.0	0.013	
K1064-06DUP	13.218	0.0	0.015	
CCV2	255.135	0.0	0.225	
CCB2	0.031	0.0	0.004	
K1064-06	12.941	0.0	0.015	
K1064-06DUP	13.079	0.0	0.015	
CCV2	257.868	0.0	0.227	
CCB2	-0.112	0.0	0.003	

N 18
Mean 50.170
SD 95.4554
CV% 190.26

AS

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

1/8/2019 12:24


Test Total CN

Accepted 1/8/2019 12:24

Factor 1153
 Bias 0.003

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.004	0.0216	0.0000	
2	5.0PPBCN	0.008	4.9423	5.0000	
3	10PPBCN	0.012	9.9495	10.0000	
4	50PPBCN	0.046	49.3601	50.0000	
5	100PPBCN	0.091	100.4535	100.0000	
6	250PPBCN	0.221	250.5963	250.0000	
7	500PPBCN	0.437	499.6767	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 08 14:20:04 2019

Tue Jan 08 15:00:29 2019

JS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.0216	µg/l	1/8/2019 12:23:51	
5.0PPBCN	A	Reactive CN	P	4.9423	µg/l	1/8/2019 12:23:52	
10PPBCN	A	Reactive CN	P	9.9495	µg/l	1/8/2019 12:23:53	
50PPBCN	A	Reactive CN	P	49.3601	µg/l	1/8/2019 12:23:54	
100PPBCN	A	Reactive CN	P	100.4535	µg/l	1/8/2019 12:23:55	
250PPBCN	A	Reactive CN	P	250.5963	µg/l	1/8/2019 12:23:56	
500PPBCN	A	Reactive CN	P	499.6767	µg/l	1/8/2019 12:23:57	
ICV1	S	Reactive CN	P	95.9443	µg/l	1/8/2019 14:20:04	
ICB1	S	Reactive CN	P	0.142	µg/l	1/8/2019 14:20:05	
CCV1	S	Reactive CN	P	242.4145	µg/l	1/8/2019 14:20:06	
CCB1	S	Reactive CN	P	-0.1263	µg/l	1/8/2019 14:20:07	
PB116240BL	S	Reactive CN	P	-0.0193	µg/l	1/8/2019 14:20:08	
K1062-02	S	Reactive CN	P	-0.1831	µg/l	1/8/2019 14:20:09	
K1064-03	S	Reactive CN	P	-0.0551	µg/l	1/8/2019 14:20:10	
K1065-02	S	Reactive CN	P	0.7354	µg/l	1/8/2019 14:20:11	
K1065-02DUP	S	Reactive CN	P	0.4994	µg/l	1/8/2019 14:20:12	
PB116243BL	S	Reactive CN	P	0.0448	µg/l	1/8/2019 14:20:13	
K1064-06	S	Reactive CN	P	11.5045	µg/l	1/8/2019 14:20:14	
K1064-06DUP	S	Reactive CN	P	13.2179	µg/l	1/8/2019 14:23:55	
CCV2	S	Reactive CN	P	255.1349	µg/l	1/8/2019 14:23:56	
CCB2	S	Reactive CN	P	0.0306	µg/l	1/8/2019 14:23:57	