



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB100208

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0003 Slope: 0.7721 Regression: 0.999982

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB100208

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH		Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
						HN03	H2SO4	Backgrnd	Color				
1	ICV	0.5	1	100	100		1.99	0.000	0.390	0.390	0.505	01/08/2019	15:03
2	ICB		1	100	100		2.23	0.000	0.000	0.000	0.000	01/08/2019	15:04
3	CCV1	0.5	1	100	100		2.37	0.000	0.390	0.390	0.505	01/08/2019	15:04
4	CCB1		1	100	100		1.64	0.000	0.000	0.000	0.000	01/08/2019	15:05
5	PB116246BL		1	2.52	100	7.33	1.88	0.000	0.000	0.000	0.000	01/08/2019	15:05
6	PB116246BS	20	1	2.51	100	7.20	2.23	0.000	0.377	0.377	0.488	01/08/2019	15:06
7	K1073-01		1	2.52	100	7.63	2.14	0.057	0.071	0.014	0.018	01/08/2019	15:06
8	K1074-01		1	2.53	100	7.12	2.32	0.022	0.022	0.000	0.000	01/08/2019	15:07
9	K1074-01DUP		1	2.50	100	7.21	1.77	0.020	0.020	0.000	0.000	01/08/2019	15:07
10	K1074-01MSPre	40	2	2.52	100	7.14	1.90	0.025	0.362	0.337	0.436	01/08/2019	15:08
11	K1074-01MS2Ins	1284	40	2.51	100	7.06	1.84	0.019	0.617	0.598	0.774	01/08/2019	15:08
12	K1074-01MS3Post	40	2	2.53	100	7.39	2.30	0.026	0.364	0.338	0.437	01/08/2019	15:09
13	K1074-03		1	2.51	100	7.55	1.69	0.067	0.068	0.001	0.001	01/08/2019	15:09
14	K1074-05		1	2.55	100	7.42	1.91	0.053	0.053	0.000	0.000	01/08/2019	15:10
15	CCV2	0.5	1	100	100		2.08	0.000	0.391	0.391	0.506	01/08/2019	15:10
16	CCB2		1	100	100		2.20	0.000	0.000	0.000	0.000	01/08/2019	15:11
17	K1074-07		1	2.50	100	7.81	1.81	0.041	0.042	0.001	0.001	01/08/2019	15:11
18	K1074-09		1	2.51	100	7.77	1.63	0.063	0.063	0.000	0.000	01/08/2019	15:12
19	K1074-11		1	2.52	100	7.06	1.55	0.066	0.066	0.000	0.000	01/08/2019	15:12
20	K1074-13		1	2.50	100	7.15	2.07	0.070	0.071	0.001	0.001	01/08/2019	15:13
21	K1074-15		1	2.53	100	7.06	2.20	0.071	0.071	0.000	0.000	01/08/2019	15:13
22	K1074-17		1	2.54	100	7.04	2.33	0.038	0.038	0.000	0.000	01/08/2019	15:14
23	K1074-19		1	2.50	100	7.32	1.87	0.055	0.055	0.000	0.000	01/08/2019	15:14
24	CCV3	0.5	1	100	100		1.56	0.000	0.393	0.393	0.509	01/08/2019	15:15
25	CCB3		1	100	100		2.33	0.000	0.000	0.000	0.000	01/08/2019	15:15