



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100211

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/08/2019	18:00
2	CAL2	1	Water	NA	NA	21.1	6.99	01/08/2019	18:02
3	CAL3	1	Water	NA	NA	21.1	10.02	01/08/2019	18:02
4	ICV	1	Water	NA	NA	21.1	7.00	01/08/2019	18:03
5	CCV1	1	Water	NA	NA	21.1	2.02	01/08/2019	18:50
6	K1086-01	1	Water	NA	NA	17.1	7.76	01/08/2019	18:51
7	K1086-01DUP	1	Water	NA	NA	17.2	7.77	01/08/2019	18:52
8	CCV2	1	Water	NA	NA	21.1	12.01	01/08/2019	19:00

WORKLIST(Hardcopy Internal Chain)

LB 100217

WorkList Name : corrosivity k1086

WorkList ID : 121116

Date : 1/8/2019 4:37:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2019	Water	K1086-01	Corrosivity	Cool 4 deg C	PSEG01	E61	CF-1819-W	01/08/2019	9040C

Date/Time 01-08-19 4:40 PM

Received by: SA

Relinquished by: SA

Date/Time

Received by:

Relinquished by:

01-08-19 6:00 PM

SA

SA