



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100219

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	26.0	78.8	79.4	01/10/2019	10:03
2	K1086-01		1	8	>100.0	>212.0	>212.0	01/10/2019	10:15
3	K1086-02		1	7	>100.0	>212.0	>212.0	01/10/2019	10:24
4	K1095-01		1	6	>100.0	>212.0	>212.0	01/10/2019	10:30
5	K1100-10		1	8	>100.0	>212.0	>212.0	01/10/2019	10:36
6	K1100-10DUP		1	8	>100.0	>212.0	>212.0	01/10/2019	10:48

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

00219

WorkList Name : FP1919

WorkList ID : 121136

Date : 1/9/2019 9:30:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2019	Water	K1086-01	Flash Point	Cool 4 deg C	PSEG01	E61	CF-1819-W	01/08/2019	1010A
01/11/2019	Water	K1086-02	Flash Point	Cool 4 deg C	PSEG01	E61	CF-1819-W	01/08/2019	1010A
01/22/2019	Water	K1095-01	Flash Point	Cool 4 deg C	DIST01	F41	2-AVE-29	01/09/2019	1010A
01/14/2019	Water	K1100-10	Flash Point	Cool 4 deg C	PSEG01	E61	WATER-DRUMS-A	01/09/2019	1010A

Date/Time 01-10-19 9:40

Received by: AD

Relinquished by: OP

Date/Time 01-10-19 11:10

Received by: AD

Relinquished by: AD