



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100221

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/09/2019	09:40
2	CAL2	1	Water	NA	NA	21.3	7.00	01/09/2019	09:41
3	CAL3	1	Water	NA	NA	21.3	10.02	01/09/2019	09:42
4	ICV	1	Water	NA	NA	21.3	7.00	01/09/2019	09:43
5	CCV1	1	Water	NA	NA	21.3	2.02	01/09/2019	09:45
6	K1004-01	1	Solid	20.03	20	23.2	10.06	01/09/2019	09:46
7	K1004-01DUP	1	Solid	20.04	20	23.2	10.07	01/09/2019	09:47
8	CCV2	1	Water	NA	NA	21.3	12.01	01/09/2019	09:48

19100221

WORKLIST(Hardcopy Internal Chain)

WorkList Name : K1004 Corrosivity

WorkList ID : 121134

Date : 1/9/2019 9:30:08 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2018	Solid	K1004-01	Corrosivity	Cool 4 deg C	ENV177	--Select--	BS01-TO-BS09-B2	12/31/2018	9045D

Date/Time 01-09-14 01:20 PM
 Received by: Sh
 Relinquished by: Sh

Date/Time 01-09-14 10:20 PM
 Received by: Sh
 Relinquished by: Sh