


LB100231

INSTRUMENT ID - CV1

7470A

Linear



A= 0.0000e+000

B= 2.4472e-004

C= 3.0089e-002

Rho= 0.9999664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-185	0.000	-185				
0.2	0.200	0.233	0.033	830	0.0 %	830				
2.5	2.500	2.479	-0.021	10006	0.0 %	10006				
5.0	5.000	5.023	0.023	20401	0.0 %	20401				
7.5	7.500	7.452	-0.048	30327	0.0 %	30327				
10.0	10.000	10.029	0.029	40860	0.0 %	40860	0			

LB100231 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jan 2019 09:36:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-185	-	0.0000	7470A	PPB	10 Jan 2019 09:37:13	S	Std
0.2 - 1	50.2	830	-	0.2000	7470A	PPB	10 Jan 2019 09:39:22	S	Std
2.5 - 1	5.25	10006	-	2.5000	7470A	PPB	10 Jan 2019 09:44:18	S	Std
5.0 - 1	55.0	20401	-	5.0000	7470A	PPB	10 Jan 2019 09:46:27	S	Std
7.5 - 1	57.5	30327	-	7.5000	7470A	PPB	10 Jan 2019 09:48:35	S	Std
10.0 - 1	510.0	40860	-	10.0000	7470A	PPB	10 Jan 2019 09:50:42	S	Std
ICV56 - 1	ICV56	15862	3.9118	-	7470A	PPB	10 Jan 2019 09:55:04	U	SMPL
ICB56 - 1	ICB56	-122	0.0002	-	7470A	PPB	10 Jan 2019 09:57:10	U	SMPL
CCV11 - 1	CCV11	20582	5.0668	-	7470A	PPB	10 Jan 2019 09:59:17	U	SMPL
CCB11 - 1	CCB11	-95	0.0068	-	7470A	PPB	10 Jan 2019 10:01:23	U	SMPL
CRA - 1	CRA	927	0.2569	-	7470A	PPB	10 Jan 2019 10:03:30	U	SMPL
HighStd - 1	HighStd	40142	9.8535	-	7470A	PPB	10 Jan 2019 10:05:37	U	SMPL
ChkStd - 1	ChkStd	28971	7.1197	-	7470A	PPB	10 Jan 2019 10:07:44	U	SMPL
PB116273BL - 1	PBW	-77	0.0112	-	7470A	PPB	10 Jan 2019 10:09:52	U	SMPL
PB116273BS - 1	LCSW	15375	3.7926	-	7470A	PPB	10 Jan 2019 10:11:59	U	SMPL
K1006-02 - 1	JC-03-010819-A	-30	0.0227	-	7470A	PPB	10 Jan 2019 10:14:07	U	SMPL
K1006-02DUP - 1	JC-03-010819-ADUP	26	0.0365	-	7470A	PPB	10 Jan 2019 10:16:15	U	SMPL
K1006-02MS - 1	JC-03-010819-AMS	13670	3.3753	-	7470A	PPB	10 Jan 2019 10:18:23	U	SMPL
K1006-02MSD - 1	JC-03-010819-AMSD	15053	3.7138	-	7470A	PPB	10 Jan 2019 10:20:32	U	SMPL
K1006-04 - 1	JC-03-010819-B	22	0.0355	-	7470A	PPB	10 Jan 2019 10:22:39	U	SMPL
CCV12 - 1	CCV12	20308	4.9998	-	7470A	PPB	10 Jan 2019 10:24:45	U	SMPL
CCB12 - 1	CCB12	-67	0.0137	-	7470A	PPB	10 Jan 2019 10:26:52	U	SMPL
K1006-06 - 1	JC-03-010819-C	252	0.0918	-	7470A	PPB	10 Jan 2019 10:29:01	U	SMPL
K1006-08 - 1	JC-03-010819-D	44	0.0409	-	7470A	PPB	10 Jan 2019 10:31:08	U	SMPL
K1006-10 - 1	JC-03-010819-E	86	0.0511	-	7470A	PPB	10 Jan 2019 10:33:14	U	SMPL
K1006-12 - 1	JC-03-010819-F	36	0.0389	-	7470A	PPB	10 Jan 2019 10:35:21	U	SMPL
K1006-14 - 1	JC-03-010819-G	39	0.0396	-	7470A	PPB	10 Jan 2019 10:37:28	U	SMPL
K1006-16 - 1	JC-03-010819-H	45	0.0411	-	7470A	PPB	10 Jan 2019 10:39:36	U	SMPL
K1006-18 - 1	JC-03-010819-I	46	0.0413	-	7470A	PPB	10 Jan 2019 10:41:44	U	SMPL
K1006-20 - 1	JC-03-010819-J	48	0.0418	-	7470A	PPB	10 Jan 2019 10:43:52	U	SMPL
K1008-02 - 1	HD-02-010819	44	0.0409	-	7470A	PPB	10 Jan 2019 10:46:01	U	SMPL
K1082-02 - 1	SP-1-1	45	0.0411	-	7470A	PPB	10 Jan 2019 10:48:08	U	SMPL
CCV13 - 1	CCV13	20547	5.0583	-	7470A	PPB	10 Jan 2019 10:50:15	U	SMPL
CCB13 - 1	CCB13	-57	0.0161	-	7470A	PPB	10 Jan 2019 10:52:22	U	SMPL
K1082-04 - 1	SP-1-2	189	0.0763	-	7470A	PPB	10 Jan 2019 10:54:29	U	SMPL
K1082-06 - 1	SP-1-3	159	0.0690	-	7470A	PPB	10 Jan 2019 10:56:37	U	SMPL
K1082-08 - 1	SP-1-4	90	0.0521	-	7470A	PPB	10 Jan 2019 10:58:44	U	SMPL
K1082-10 - 1	SP-1-5	52	0.0428	-	7470A	PPB	10 Jan 2019 11:00:52	U	SMPL
K1082-12 - 1	SP-1-6	55	0.0435	-	7470A	PPB	10 Jan 2019 11:02:59	U	SMPL
K1087-04 - 1	MH-448-C	54	0.0433	-	7470A	PPB	10 Jan 2019 11:05:07	U	SMPL
K1087-08 - 1	MH-447-C	47	0.0416	-	7470A	PPB	10 Jan 2019 11:07:15	U	SMPL
K1087-12 - 1	MH-445	50	0.0423	-	7470A	PPB	10 Jan 2019 11:09:24	U	SMPL
PB116273TB - 1	PB116247TB	84	0.0506	-	7470A	PPB	10 Jan 2019 11:11:33	U	SMPL
CCV14 - 1	CCV14	20483	5.0426	-	7470A	PPB	10 Jan 2019 11:13:40	U	SMPL
CCB14 - 1	CCB14	-67	0.0137	-	7470A	PPB	10 Jan 2019 11:15:48	U	SMPL
K1006-02LX5 - 1		19	0.0347	-	7470A	PPB	10 Jan 2019 11:17:55	U	SMPL
K1006-02A - 1		16045	3.9565	-	7470A	PPB	10 Jan 2019 11:20:02	U	SMPL
CCV15 - 1	CCV15	20035	4.9330	-	7470A	PPB	10 Jan 2019 11:22:09	U	SMPL
CCB15 - 1	CCB15	-81	0.0103	-	7470A	PPB	10 Jan 2019 11:24:16	U	SMPL