



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100239

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/10/2019	12:10
2	CAL2	1	Water	NA	NA	21.1	7.01	01/10/2019	12:11
3	CAL3	1	Water	NA	NA	21.1	10.03	01/10/2019	12:12
4	ICV	1	Water	NA	NA	21.1	6.99	01/10/2019	12:13
5	CCV1	1	Water	NA	NA	21.1	2.02	01/10/2019	12:45
6	K1086-02	1	Water	NA	NA	17.1	7.76	01/10/2019	12:46
7	K1086-02DUP	1	Water	NA	NA	17.1	7.77	01/10/2019	12:47
8	K1100-10	1	Water	NA	NA	14.5	9.72	01/10/2019	12:48
9	CCV2	1	Water	NA	NA	21.1	12.02	01/10/2019	12:49

13100239

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k10086

WorkList ID : 121207

Date : 1/10/2019 12:01:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2019	Water	K1086-02	Corrosivity	Cool 4 deg C	PSEG01	E61	CF-1819-W	01/08/2019	9040C
01/09/2019	Water	K1100-10	Corrosivity	Cool 4 deg C	PSEG01	E61	WATER-DRUMS-A	01/09/2019	9040C

Date/Time 6-1-10-19 12:10 PM
Received by: SO
Relinquished by: CA

Date/Time 01-10-19 21:00 PM
Received by: CD
Relinquished by: JP