



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100241

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/10/2019	12:10
2	CAL2	1	Water	NA	NA	21.1	7.01	01/10/2019	12:11
3	CAL3	1	Water	NA	NA	21.1	10.03	01/10/2019	12:12
4	ICV	1	Water	NA	NA	21.1	6.99	01/10/2019	12:13
5	CCV1	1	Water	NA	NA	21.2	2.02	01/10/2019	12:50
6	K1090-01	1	Solid	20.03	20	19.7	10.37	01/10/2019	13:00
7	K1090-01DUP	1	Solid	20.04	20	19.8	10.38	01/10/2019	13:01
8	K1090-02	1	Solid	20.02	20	22.5	9.71	01/10/2019	13:02
9	K1090-03	1	Solid	20.04	20	20.5	10.17	01/10/2019	13:03
10	K1090-04	1	Solid	20.02	20	21.4	9.81	01/10/2019	13:05
11	K1090-05	1	Solid	20.03	20	20.7	9.72	01/10/2019	13:06
12	K1090-06	1	Solid	20.04	20	20.6	9.86	01/10/2019	13:09
13	K1090-07	1	Solid	20.02	20	20.6	9.73	01/10/2019	13:10
14	K1090-08	1	Solid	20.02	20	18.2	10.60	01/10/2019	13:11
15	K1090-09	1	Solid	20.03	20	18.5	10.60	01/10/2019	13:12
16	CCV2	1	Water	NA	NA	21.1	12.01	01/10/2019	13:13
17	K1090-10	1	Solid	20.04	20	22.1	8.78	01/10/2019	13:15
18	CCV3	1	Water	NA	NA	21.2	2.02	01/10/2019	13:16

28100241

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph K1090

WorkList ID : 121209

Date : 1/10/2019 12:03:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2018	Solid	K1090-01	pH	Cool 4 deg C	ENVI77	--Select--	BS-01	12/03/2018	9045D
12/03/2018	Solid	K1090-02	pH	Cool 4 deg C	ENVI77	--Select--	BS-02	12/03/2018	9045D
12/04/2018	Solid	K1090-03	pH	Cool 4 deg C	ENVI77	--Select--	BS-03	12/04/2018	9045D
12/03/2018	Solid	K1090-04	pH	Cool 4 deg C	ENVI77	--Select--	BS-04	12/03/2018	9045D
12/03/2018	Solid	K1090-05	pH	Cool 4 deg C	ENVI77	--Select--	BS-05	12/03/2018	9045D
12/04/2018	Solid	K1090-06	pH	Cool 4 deg C	ENVI77	--Select--	BS-06	12/04/2018	9045D
12/04/2018	Solid	K1090-07	pH	Cool 4 deg C	ENVI77	--Select--	BS-07	12/04/2018	9045D
12/04/2018	Solid	K1090-08	pH	Cool 4 deg C	ENVI77	--Select--	BS-08	12/04/2018	9045D
12/06/2018	Solid	K1090-09	pH	Cool 4 deg C	ENVI77	--Select--	BS-09	12/06/2018	9045D
11/29/2018	Solid	K1090-10	pH	Cool 4 deg C	ENVI77	--Select--	B2	11/29/2018	9045D

Date/Time 12-10-19 12:11:00 PM

Received by: JS

Relinquished by: _____

Date/Time 01-10-19 21:00 PM

Received by: JSRelinquished by: JS