

LB100247


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

62392



A= 0.0000e+000

B= 1.6174e-004

C= -5.8155e-002

Rho= 0.9998971

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	280	0.000	280				
0.2	0.200	0.184	-0.016	1498	0.0 %	1498				
2.5	2.500	2.529	0.029	15995	0.0 %	15995				
5.0	5.000	5.067	0.067	31688	0.0 %	31688				
7.5	7.500	7.400	-0.100	46113	0.0 %	46113				
10.0	10.000	10.033	0.033	62392	0.0 %	62392				

LB100247 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jan 2019 14:11:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	280	-	0.0000	7471B	PPB	10 Jan 2019 14:13:32	S	Std
0.2 - 1	50.2	1498	-	0.2000	7471B	PPB	10 Jan 2019 14:15:50	S	Std
2.5 - 1	52.5	15995	-	2.5000	7471B	PPB	10 Jan 2019 14:18:08	S	Std
5.0 - 1	55.0	31688	-	5.0000	7471B	PPB	10 Jan 2019 14:20:26	S	Std
7.5 - 1	57.5	46113	-	7.5000	7471B	PPB	10 Jan 2019 14:22:43	S	Std
10.0 - 1	510.0	62392	-	10.0000	7471B	PPB	10 Jan 2019 14:35:03	S	Std
ICV57 - 1	ICV57	23442	3.7333	-	7471B	PPB	10 Jan 2019 14:37:52	U	SMPL
ICB57 - 1	ICB57	156	-0.0329	-	7471B	PPB	10 Jan 2019 14:40:08	U	SMPL
CCV16 - 1	CCV16	30516	4.8774	-	7471B	PPB	10 Jan 2019 14:42:24	U	SMPL
CCB16 - 1	CCB16	144	-0.0349	-	7471B	PPB	10 Jan 2019 14:44:40	U	SMPL
CRA - 1	CRA	1431	0.1733	-	7471B	PPB	10 Jan 2019 14:46:57	U	SMPL
HighStd - 1	HighStd	63115	10.1499	-	7471B	PPB	10 Jan 2019 14:49:13	U	SMPL
ChkStd - 1	ChkStd	41534	6.6594	-	7471B	PPB	10 Jan 2019 14:51:30	U	SMPL
PB116290BL - 1	PBS	125	-0.0379	-	7471B	PPB	10 Jan 2019 14:53:49	U	SMPL
PB116290BS - 1	LCSS	21096	3.3538	-	7471B	PPB	10 Jan 2019 14:56:07	U	SMPL
K1005-01 - 1	HR-06-010918-A	5617	0.8503	-	7471B	PPB	10 Jan 2019 14:58:24	U	SMPL
K1005-03 - 1	HR-06-010918-B	37131	5.9473	-	7471B	PPB	10 Jan 2019 15:00:42	U	SMPL
K1005-05 - 1	HR-06-010918-C	87712	14.1281	-	7471B	PPB	10 Jan 2019 15:02:59	U	SMPL
K1015-01 - 1	SU-03-010918-A	1203	0.1364	-	7471B	PPB	10 Jan 2019 15:05:18	U	SMPL
K1015-03 - 1	SU-03-010918-B	5622	0.8511	-	7471B	PPB	10 Jan 2019 15:07:37	U	SMPL
CCV17 - 1	CCV17	30266	4.8370	-	7471B	PPB	10 Jan 2019 15:09:53	U	SMPL
CCB17 - 1	CCB17	121	-0.0386	-	7471B	PPB	10 Jan 2019 15:12:09	U	SMPL
K1087-14 - 1	MH-444-C	3216	0.4620	-	7471B	PPB	10 Jan 2019 15:14:26	U	SMPL
K1094-01 - 1	CROSSWICKS-DRUM-1	531	0.0277	-	7471B	PPB	10 Jan 2019 15:16:42	U	SMPL
K1094-03 - 1	CROSSWICKS-DRUM-2	1565	0.1950	-	7471B	PPB	10 Jan 2019 15:18:58	U	SMPL
K1096-01 - 1	390103023	218004	35.2011	-	7471B	PPB	10 Jan 2019 15:21:15	U	SMPL
K1096-02 - 1	390103027.1	826	0.0754	-	7471B	PPB	10 Jan 2019 15:23:31	U	SMPL
K1099-01 - 1	TP-15	2866	0.4054	-	7471B	PPB	10 Jan 2019 15:25:54	U	SMPL
K1099-01DUP - 1	TP-15DUP	2807	0.3958	-	7471B	PPB	10 Jan 2019 15:28:11	U	SMPL
K1099-01MS - 1	TP-15MS	24348	3.8798	-	7471B	PPB	10 Jan 2019 15:30:29	U	SMPL
K1099-01MSD - 1	TP-15MSD	23853	3.7998	-	7471B	PPB	10 Jan 2019 15:32:47	U	SMPL
K1099-05 - 1	TP-16	10180	1.5883	-	7471B	PPB	10 Jan 2019 15:35:03	U	SMPL
CCV18 - 1	CCV18	30594	4.8900	-	7471B	PPB	10 Jan 2019 15:37:21	U	SMPL
CCB18 - 1	CCB18	111	-0.0402	-	7471B	PPB	10 Jan 2019 15:39:37	U	SMPL
K1099-09 - 1	TP-17-14	142426	22.9774	-	7471B	PPB	10 Jan 2019 15:41:54	U	SMPL
K1099-13 - 1	TP-18	13704	2.1583	-	7471B	PPB	10 Jan 2019 15:44:11	U	SMPL
K1099-17 - 1	TP-19-20	10047	1.5668	-	7471B	PPB	10 Jan 2019 15:46:30	U	SMPL
K1099-21 - 1	TP-21	87947	14.1661	-	7471B	PPB	10 Jan 2019 15:48:47	U	SMPL
K1100-01 - 1	SOIL-DRUM-A	213	-0.0237	-	7471B	PPB	10 Jan 2019 15:51:04	U	SMPL
K1100-03 - 1	SOIL-DRUM-B	429	0.0112	-	7471B	PPB	10 Jan 2019 15:53:24	U	SMPL
K1100-05 - 1	SOIL-DRUM-C	1260	0.1456	-	7471B	PPB	10 Jan 2019 15:55:41	U	SMPL
K1100-07 - 1	SOIL-DRUM-D	2730	0.3834	-	7471B	PPB	10 Jan 2019 15:57:59	U	SMPL
K1099-01LX5 - 1		730	0.0599	-	7471B	PPB	10 Jan 2019 16:00:16	U	SMPL
K1099-01A - 1		26698	4.2599	-	7471B	PPB	10 Jan 2019 16:02:33	U	SMPL
CCV19 - 1	CCV19	30826	4.9275	-	7471B	PPB	10 Jan 2019 16:04:49	U	SMPL
CCB19 - 1	CCB19	132	-0.0368	-	7471B	PPB	10 Jan 2019 16:07:06	U	SMPL
K1005-05DLX2 - 1		45519	7.3039	-	7471B	PPB	10 Jan 2019 16:09:24	U	SMPL
K1096-01DLX5 - 1		48525	7.7901	-	7471B	PPB	10 Jan 2019 16:11:40	U	SMPL
K1099-09DLX5 - 1		29546	4.7205	-	7471B	PPB	10 Jan 2019 16:13:58	U	SMPL
K1099-21DLX2 - 1		45376	7.2808	-	7471B	PPB	10 Jan 2019 16:16:17	U	SMPL
CCV20 - 1	CCV20	30662	4.9010	-	7471B	PPB	10 Jan 2019 16:18:35	U	SMPL
CCB20 - 1	CCB20	120	-0.0387	-	7471B	PPB	10 Jan 2019 16:20:54	U	SMPL