

213100248

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/10/2019 15:50

Test: Phenolics-

Sample Id	Result <i>mg/d</i>	Dil. 1 +	Response	Errors
ICV1	1.028	0.0	0.085	
ICB1	0.001	0.0	0.007	
CCV1	0.984	0.0	0.082	
CCB1	-0.002	0.0	0.007	
PB116302BL	-0.003	0.0	0.007	
PB116302BS	1.022	0.0	0.084	
K1018-01	0.004	0.0	0.008	
K1095-01	0.055	0.0	0.012	
K1095-01DUP	0.052	0.0	0.011	
K1095-01MS	1.055	0.0	0.087	
K1095-01MSD	1.087	0.0	0.089	
K1018-02	0.060	0.0	0.012	
CCV2	0.996	0.0	0.082	
CCB2	-0.000	0.0	0.007	

N	14
Mean	0.453
SD	0.5186
CV%	114.55

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 10 14:57:09 2019

Thu Jan 10 15:41:24 2019

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0492	mg/l	1/10/2019 14:57:09	
0.1PPM	A	Phenolics-	P	0.1018	mg/l	1/10/2019 14:57:10	
0.25PPM	A	Phenolics-	P	0.2483	mg/l	1/10/2019 14:57:11	
0.50PPM	A	Phenolics-	P	0.5101	mg/l	1/10/2019 14:57:12	
1.0PPM	A	Phenolics-	P	0.986	mg/l	1/10/2019 14:57:13	
2.0PPM	A	Phenolics-	P	2.0046	mg/l	1/10/2019 14:57:14	
ICV1	S	Phenolics-	P	1.0283	mg/l	1/10/2019 15:37:25	
ICB1	S	Phenolics-	P	0.0008	mg/l	1/10/2019 15:37:26	
CCV1	S	Phenolics-	P	0.9839	mg/l	1/10/2019 15:37:27	
CCB1	S	Phenolics-	P	-0.0018	mg/l	1/10/2019 15:37:28	
PB116302BL	S	Phenolics-	P	-0.003	mg/l	1/10/2019 15:37:29	
PB116302BS	S	Phenolics-	P	1.0223	mg/l	1/10/2019 15:37:30	
K1018-01	S	Phenolics-	P	0.0035	mg/l	1/10/2019 15:37:31	
K1095-01	S	Phenolics-	P	0.0552	mg/l	1/10/2019 15:37:32	
K1095-01DUP	S	Phenolics-	P	0.0518	mg/l	1/10/2019 15:37:33	
K1095-01MS	S	Phenolics-	P	1.0546	mg/l	1/10/2019 15:37:34	
K1095-01MSD	S	Phenolics-	P	1.0872	mg/l	1/10/2019 15:37:35	
K1018-02	S	Phenolics-	P	0.0595	mg/l	1/10/2019 15:37:36	
CCV2	S	Phenolics-	P	0.9959	mg/l	1/10/2019 15:41:23	
CCB2	S	Phenolics-	P	-0.0001	mg/l	1/10/2019 15:41:24	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

1/10/2019 14:58


Test Phenolics-

Accepted 1/10/2019 14:58

Factor 13.26
 Bias 0.007

Coeff. of det. 0.999883

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.011	0.0492	0.0500	
2	Phenol-2	0.015	0.1018	0.1000	
3	Phenol-2	0.026	0.2483	0.2500	
4	Phenol-2	0.046	0.5101	0.5000	
5	Phenol-2	0.082	0.9860	1.0000	
6	Phenol-2	0.159	2.0046	2.0000	